

DOG1

Concentrated and Prediluted Monoclonal Antibody
901-385-082719

Catalog Number:	CM 385 A, C	PM 385 AA
Description:	0.1, 1.0 mL, conc.	6.0 mL, RTU
Dilution:	1:100	Ready-to-use
Diluent:	Da Vinci Green	N/A

Intended Use:

For In Vitro Diagnostic Use

DOG1 [DOG1.1] is a mouse monoclonal antibody that is intended for laboratory use in the qualitative identification of DOG1 protein by immunohistochemistry (IHC) in formalin-fixed paraffin-embedded (FFPE) human tissues. The clinical interpretation of any staining or its absence should be complemented by morphological studies using proper controls and should be evaluated within the context of the patient's clinical history and other diagnostic tests by a qualified pathologist.

Summary and Explanation:

DOG1 is a cell surface protein of unknown function selectively expressed in gastrointestinal stromal tumors (GIST). DOG1 expression has been reported to be a very sensitive and specific marker for GIST in paraffin-embedded tissue. Studies have shown that among GIST cases with KIT mutations, DOG1 detected 11% more cases than CD117. In KIT/CD117 negative and PDGFRA-mutant GIST cases, DOG1 increased the accuracy of GIST diagnosis (2). As a result of its localization in the cell membrane, its absence in the majority of normal tissue and its presence in most GIST tissue, DOG1 may be an additional target in the diagnosis and treatment of GIST (1-4). DOG1 immunoreactivity is seen in fewer cases of mesenchymal, epithelial tumors, seminomas and melanomas when compared with CD117.

Principle of Procedure:

Antigen detection in tissues and cells is a multi-step immunohistochemical process. The initial step binds the primary antibody to its specific epitope. After labeling the antigen with a primary antibody, a one-, two- or three-step detection procedure can be employed. The one-step procedure will feature an enzyme-labeled polymer that binds to the primary antibody. A two-step procedure will feature a secondary antibody added to bind to the primary antibody. An enzyme-labeled polymer is then added to bind to the secondary antibody. The three-step detection procedure will feature a secondary antibody added to bind to the primary antibody followed by a linker antibody step for maximum binding. An enzyme-labeled polymer is then added to bind to the linker antibody. These detections of the bound antibodies are evidenced by a colorimetric reaction.

Source: Mouse monoclonal

Species Reactivity: Human

Clone: DOG1.1

Isotype: IgG1/kappa

Protein Concentration: Call for lot specific Ig concentration.

Epitope/Antigen: Synthetic peptide of human DOG1

Cellular Localization: Membrane & cytoplasmic

Positive Tissue Control: Gastrointestinal stromal tumors

Known Applications:

Immunohistochemistry (formalin-fixed paraffin-embedded tissues)

Supplied As: Buffer with protein carrier and preservative

Storage and Stability:

Store at 2°C to 8°C. The product is stable to the expiration date printed on the label, when stored under these conditions. Do not use after expiration date. Diluted reagents should be used promptly; any remaining reagent should be stored at 2°C to 8°C.

GAR
CAM

BIOCARE
BIOCARE MEDICAL

1-800-799-9499

CE IVD

DOG1

REF CM385A
LOT 101421A

2023/10
0.1 ml
REV. A

OAI 385 T60

60 tests, RTU

Ready-to-use

N/A

N/A

Protocol Recommendations (VALENT® Automated Slide Staining Platform):

VLT385 is intended for use with the VALENT. Refer to the User Manual for specific instructions for use. Protocol parameters in the Protocol Manager should be programmed as follows:

Deparaffinization: Deparaffinize for 8 minutes with Val DePar.

Pretreatment: Perform heat retrieval at 98°C for 60 minutes using Val AR-Hi pH, 5X (use at 1X).

Peroxidase Block: Block for 5 minutes with Val Peroxidase Block.

Protein Block: Incubate for 10 minutes at RT with Val Background Block.

Primary Antibody: Incubate for 45 minutes.

Secondary: Incubate for 10 minutes with Val Mouse Secondary.

Linker: Incubate for 10 minutes with Val Universal Linker.

Polymer: Incubate for 10 minutes with Val Universal Polymer.

Chromogen: Incubate for 5 minutes with Val DAB.

Counterstain: Counterstain for 5 minutes with Val Hematoxylin.

Protocol Recommendations (intelliPATH FLX® and manual use):

Peroxide Block: Block for 5 minutes with Peroxidized 1.

Pretreatment: Perform heat retrieval using Borg or Reveal Decloaker. Refer to the Borg or Reveal Decloaker product data sheet for specific instructions.

Protein Block (Optional): Incubate for 5-10 minutes at RT with Background Punisher.

Primary Antibody: Incubate for 30-45 minutes at RT.

Probe: Incubate for 10 minutes at RT with a secondary probe.

Polymer: Incubate for 10-20 minutes at RT with a tertiary polymer.

Chromogen: Incubate for 5 minutes at RT with Biocare's DAB - OR - Incubate for 5-7 minutes at RT with Warp Red.

Counterstain: Counterstain with hematoxylin. Rinse with deionized water. Apply Tacha's Bluing Solution for 1 minute. Rinse with deionized water.

Technical Note:

This antibody, for intelliPATH FLX and manual use, has been standardized with MACH 4 detection system. Use TBS for washing steps.

Protocol Recommendations (ONCORE™ Automated Slide Staining System):

OAI385 is intended for use with the ONCORE. Refer to the User Manual for specific instructions for use. Protocol parameters in the ONCORE Automated Slide Stainer Protocol Editor should be programmed as follows:

Protocol Name: DOG1

Protocol Template (Description): Ms HRP Template 1

Dewaxing (DS Option): DS2

Antigen Retrieval (AR Option): AR1, high pH; 103°C

Reagent Name, Time, Temp.: DOG1, 30 min., 25°C

Limitations:

The optimum antibody dilution and protocols for a specific application can vary. These include, but are not limited to fixation, heat-retrieval method, incubation times, tissue section thickness and detection kit used. Due to the superior sensitivity of these unique reagents, the

Biocare Medical
60 Berry Drive
Pacheco, CA 94553
USA

IVD CE

Rev: 062117

Tel: 800-799-9499 | www.biocare.net | Fax: 925-603-8080

EC REP EMERGO EUROPE
Prinsessegracht 20
2514 AP The Hague
The Netherlands

DOG1

Concentrated and Prediluted Monoclonal Antibody
901-385-082719

BIOCARE
MEDICAL

Limitations Cont'd:

recommended incubation times and titers listed are not applicable to other detection systems, as results may vary. The data sheet recommendations and protocols are based on exclusive use of Biocare products. Ultimately, it is the responsibility of the investigator to determine optimal conditions.

Quality Control:

Refer to CLSI Quality Standards for Design and Implementation of Immunohistochemistry Assays; Approved Guideline-Second edition (I/LA28-A2) CLSI Wayne, PA, USA (www.clsi.org). 2011

Precautions:


1. This antibody contains less than 0.1% sodium azide. Concentrations less than 0.1% are not reportable hazardous materials according to U.S. 29 CFR 1910.1200, OSHA Hazard communication and EC Directive 91/155/EC. Sodium azide (NaN_3) used as a preservative is toxic if ingested. Sodium azide may react with lead and copper plumbing to form highly explosive metal azides. Upon disposal, flush with large volumes of water to prevent azide build-up in plumbing. (Center for Disease Control, 1976, National Institute of Occupational Safety and Health, 1976) (5)
2. Specimens, before and after fixation, and all materials exposed to them should be handled as if capable of transmitting infection and disposed of with proper precautions. Never pipette reagents by mouth and avoid contacting the skin and mucous membranes with reagents and specimens. If reagents or specimens come in contact with sensitive areas, wash with copious amounts of water. (6)
3. Microbial contamination of reagents may result in an increase in nonspecific staining.
4. Incubation times or temperatures other than those specified may give erroneous results. The user must validate any such change.
5. Do not use reagent after the expiration date printed on the vial.
6. The SDS is available upon request and is located at <http://biocare.net>.

Troubleshooting:

Follow the antibody specific protocol recommendations according to data sheet provided. If atypical results occur, contact Biocare's Technical Support at 1-800-542-2002.

References:

1. Espinosa I, *et al.* A novel monoclonal antibody against DOG1 is a sensitive and specific marker for gastrointestinal stromal tumors. *Am J Surg Pathol.* 2008 Feb;32 (2):210-8.
2. Miwa S, *et al.* Mutation assay of the novel gene DOG1 in gastrointestinal stromal tumors (GISTs). *J Gastroenterol.* 2008; 43(7):531-7.
3. Parfitt JR, *et al.* Gastrointestinal Kaposi's sarcoma: CD117 expression and the potential for misdiagnosis as gastrointestinal stromal tumor. *Histopathology.* 2008 Jun;52(7):816-23.
4. West RB, *et al.* The novel marker, DOG1, is expressed ubiquitously in gastrointestinal stromal tumors irrespective of KIT or PDGFRA mutation status. *Am J Pathol.* 2004 Jul;165(1):107-13.
5. Center for Disease Control Manual. Guide: Safety Management, NO. CDC-22, Atlanta, GA. April 30, 1976 "Decontamination of Laboratory Sink Drains to Remove Azide Salts."
6. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections; Approved Guideline-Fourth Edition CLSI document M29-A4 Wayne, PA 2014.



60 Berry Drive
Pacheco, CA 94553
USA



Rev: 062117

Tel: 800-799-9499 | www.biocare.net | Fax: 925-603-8080

EC REP EMERGO EUROPE
Prinsessegracht 20
2514 AP The Hague
The Netherlands