


Complain Claim Form					
Company Name	Altamedica Inc.				
Instrument Model	C-61	Installation Date			
Serial Number	601831087 IE	Problem date	22/12/2016		
Current Status	<input checked="" type="checkbox"/> Solved	<input type="checkbox"/> Unsolved	Picture or video	<input checked="" type="checkbox"/> Yes	<input type="checkbox"/> No
Problem Detailed Description	Fallen filament of the spare halogen lamp				
Maintenance Procedure & Action Taken					
Broken spares:					
Part No.	Spare Part Name	QTY	Explanation of the broken (detailed checking process)		
	Halogen Lamp	1	Fallen filament of the spare halogen lamp		
Spare parts picture					
					
Your Remark/Suggestion/Comments on the problem:					
Date : 22-Dec-16					