Gauze

Product / Application	BIS	ISO	ASTM	DIN	Other Standards
Absorbent Gauze	IS 758,	ISO 13485; ISO 9073-18; ISO 14079 (Surgical dressings — absorbent gauze, where applicable),ISO 13937; ISO 3801	ASTM D5034; ASTM D1777,ASTM D1117; ASTM D2261; ASTM D3776	DIN EN 14079 (Non-active surgical implants — absorbent gauze/dressings)	-
Gauze Roll	IS 8588,IS 758	ISO 14079 (Surgical dressings — gauze dressings),ISO 13937; ISO 3801	ASTM D5034,ASTM D1117; ASTM D2261; ASTM D3776	DIN EN 14079	-
Gauze Swab	IS 8588,IS 758	ISO 14079,ISO 13937; ISO 3801	ASTM D5034,ASTM D1117; ASTM D2261; ASTM D3776	DIN EN 14079	-
Sterile Gauze	IS 8588,IS 758	ISO 14079; ISO 11135 (EO sterilization); ISO 11137 (Radiation sterilization),ISO 13937; ISO 3801	ASTM F88; ASTM F1980; ASTM F2096 (Gross Leak Detection),ASTM D1117; ASTM D2261; ASTM D3776; ASTM D5034	DIN EN 14079; DIN EN ISO 11135 / DIN EN ISO 11137 depending on sterilization method	AAMI ST79
Non-Sterile Gauze	IS 8588,IS 758:2023	ISO 14079,ISO 13937; ISO 3801	ASTM D5034,ASTM D1117; ASTM D2261; ASTM D3776	DIN EN 14079	-
Paraffin Gauze Dressing	IS 758	ISO 14079; ISO 10993 series,ISO 13937; ISO 3801	ASTM D5034; ASTM F88 (Packaging integrity where sterile),ASTM D1117; ASTM D2261; ASTM D3776	DIN EN 14079	-
X-ray Detectable Gauze	IS 8588,IS 16950:2018; IS 758	ISO 14079; ISO 10993 series,ISO 13937; ISO 3801	ASTM F640 (Radiopacity Testing of Medical Devices); ASTM D5034,ASTM D1117; ASTM D2261; ASTM D3776	DIN EN 14079; DIN EN ISO 10993 series where applicable	-
Gauze Sponge	IS 8588,IS 758	ISO 14079,ISO 13937; ISO 3801	ASTM D5034,ASTM D1117; ASTM D2261; ASTM D3776	DIN EN 14079	-
Laparotomy Sponge	IS 8588,IS 758	ISO 14079,ISO 13937; ISO 3801	ASTM F640; ASTM D5034,ASTM D1117; ASTM D2261; ASTM D3776	DIN EN 14079	-
Gauze Pad	IS 8588,IS 758	ISO 14079,ISO 13937; ISO 3801	ASTM D5034,ASTM D1117; ASTM D2261; ASTM D3776	DIN EN 14079	-

Source field: Medical devices