Medical Packaging

Product / Application	BIS	ISO	ASTM	DIN	Other Standards
Medical Device Pouch	IS/ISO 11607-1:2019,IS 16252	ISO 11607-1 (Packaging for terminally sterilized medical devices — Materials, sterile barrier systems); ISO 11607-2 (Packaging process validation),	ASTM F88 (Seal strength of flexible barrier materials); ASTM F1929 (Dye penetration leak testing); ASTM F2096 (Gross leak testing); ASTM F1980 (Accelerated aging),ASTM D4169; ASTM F1307; ASTM F2097; ASTM F2825	DIN EN ISO 11607-1 (Packaging for terminally sterilized medical devices — Materials and sterile barrier systems); DIN EN ISO 11607-2 (Packaging process validation)	AAMI ST72 (Sterile barrier packaging validation guidance); AAMI ST79 (Sterile processing practices where applicable)
Medical Device Tray	IS/ISO 11607-1:2019,IS 16252	ISO 11607-1; ISO 11607-2,	ASTM F88; ASTM F1929; ASTM F2096; ASTM D4169 (Distribution testing),ASTM F1307; ASTM F1980; ASTM F2097; ASTM F2825	DIN EN ISO 11607-1; DIN EN ISO 11607-2	AAMI ST77 (Containment devices for reusable medical devices); AAMI ST79
Thermoformed Medical Tray	IS/ISO 11607-1:2019,IS 16252	ISO 11607-1; ISO 11607-2,	ASTM F88; ASTM F2096; ASTM D6241 (Compressive strength of packaging where applicable),ASTM D4169; ASTM F1307; ASTM F1980; ASTM F2097; ASTM F2825	DIN EN ISO 11607-1; DIN EN ISO 11607-2	AAMI ST77; AAMI ST79
Medical Blister Pack	IS/ISO 11607-1:2019,IS 11356; IS 2508	ISO 11607-1; ISO 11607-2,ISO 11469; ISO 11833	ASTM F88; ASTM F1929; ASTM F2096,ASTM D3078; ASTM F1249; ASTM F1927; ASTM F2029; ASTM F3169	DIN EN ISO 11607-1; DIN EN ISO 11607-2	AAMI ST72
Medical Clamshell Pack	IS/ISO 11607-1:2019,IS 16252	ISO 11607-1,	ASTM F88; ASTM F1929,ASTM D4169; ASTM F1307; ASTM F1980; ASTM F2097; ASTM F2825	DIN EN ISO 11607-1	AAMI ST72
Medical Peelable Pouch	IS/ISO 11607-1:2019,IS 16252	ISO 11607-1; ISO 11607-2,	ASTM F88 (Peelable seal strength); ASTM F1929,ASTM D4169; ASTM F1307; ASTM F1980; ASTM F2097; ASTM F2825	DIN EN ISO 11607-1; DIN EN ISO 11607-2	AAMI ST72
Tyvek® Medical Packaging	IS/ISO 11607-1:2019,IS 16252	ISO 11607-1; ISO 11607-2,	ASTM F88; ASTM F1929; ASTM F2096,ASTM D4169; ASTM F1307; ASTM F1980; ASTM F2097; ASTM F2825	DIN EN ISO 11607-1; DIN EN ISO 11607-2	AAMI ST72
Sterile Barrier Packaging	IS/ISO 11607-1:2019,IS 16252	ISO 11607-1; ISO 11607-2,	ASTM F88; ASTM F1929; ASTM F2096; ASTM F1980,ASTM D4169; ASTM F1307; ASTM F2097; ASTM F2825	DIN EN ISO 11607-1; DIN EN ISO 11607-2	AAMI ST72 (Sterile barrier systems for terminally sterilized medical devices)
Medical Device Carton	IS/ISO 11607-1:2019*,IS 16252	ISO 11607-1 (Packaging for terminally sterilized medical devices),	ASTM D6251 (Wood-Cleated Shipping Boxes where applicable); ASTM D6256 (Rigid box performance); ASTM D4169,ASTM F1307; ASTM F1980; ASTM F2097; ASTM F2825	DIN EN ISO 11607-1; DIN EN 868 series (Packaging materials and systems for sterilized medical devices)	-
Medical Device Shipping Box	IS 16252	ISO 11607-1; ISO 15278 (Packaging requirements, applicable),	ASTM D4169 (Performance testing of shipping containers); ASTM D6251/D6256 where applicable,ASTM F1307; ASTM F1980; ASTM F2097; ASTM F2825	DIN EN ISO 11607-1; DIN EN 868 series	-
Medical Device Label	IS 16252	ISO 15223-1 (Symbols for medical device labelling); ISO 20417 (Information supplied by manufacturer),ISO 11607	ASTM D5486 (Pressure-sensitive labels for shipping containers); ASTM F2250 (Coating/label testing where applicable),ASTM D4169; ASTM F1307; ASTM F1980; ASTM F2097; ASTM F2825	DIN EN ISO 15223-1 (Symbols for medical device labelling); DIN EN ISO 20417 (Information supplied by manufacturer)	AAMI HE75 (Human factors engineering for medical devices where applicable)
Medical Device Container	IS/ISO 11607-1:2019*,IS 16252	ISO 11607-1; ISO 11607-2,	ASTM D4169; ASTM F2096; ASTM F1980,ASTM F1307; ASTM F2097;	DIN EN ISO 11607-1; DIN EN ISO 11607-2	AAMI ST77 (Containment devices); AAMI ST79

Product / Application	BIS	ISO	ASTM	DIN	Other Standards
			ASTM F2825		
Specimen Transport Container	IS 16252	ISO 15189 (Laboratory quality requirements); ISO 6710 (Specimen collection systems, applicable),ISO 11607	ASTM D4169; ASTM F2096,ASTM F1307; ASTM F1980; ASTM F2097; ASTM F2825	DIN EN ISO 15189 (Medical laboratories — Requirements for quality and competence); DIN EN ISO 6710 (Single-use containers for venous blood specimen collection where applicable)	-
Biohazard Transport Container	IS 16252	ISO 15190 (Medical laboratories — Safety requirements); ISO 15189,ISO 11607	ASTM D4169; ASTM D6251/D6256 where applicable,ASTM F1307; ASTM F1980; ASTM F2097; ASTM F2825	DIN EN ISO 15190 (Medical laboratories — Requirements for safety); DIN EN ISO 15189	-
Pharmaceutical-Medical Combination Packaging	IS/ISO 11607-1:2019*,IS 10146; IS 10910	ISO 11607-1; ISO 15378 (Primary packaging materials for medicinal products),ISO 9001	ASTM F88; ASTM F1929; ASTM F2096; ASTM F1980,ASTM D4169; ASTM F1249; ASTM F1307; ASTM F1927; ASTM F2029	DIN EN ISO 11607-1; DIN EN ISO 15378 (Primary packaging materials for medicinal products)	AAMI TIR57 (Principles for combination products where applicable)

Source field: Medical devices