Needles

Product / Application	BIS	ISO	ASTM	DIN	Other Standards
Hypodermic Needle	IS 10654,	ISO 9626 (Stainless steel needle tubing); ISO 7864 (Sterile hypodermic needles for single use),	ASTM F640 (Radiopacity of medical devices); ASTM F1841 (Coated medical device lubricity where applicable),ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 7864 (Sterile hypodermic needles for single use); DIN EN ISO 9626 (Needle tubing)	AAMI/ANSI ST72 (Sterile packaging where applicable)
Sterile Disposable Needle	IS 10654,	ISO 7864,ISO 9626	ASTM F640; ASTM F88 (sterile packaging seal strength),ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 7864	AAMI ST72
Spinal Needle	IS 7350,IS 10654	ISO 9626; ISO 7864,	ASTM F640; ASTM F1841,ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 7864; DIN EN ISO 9626	-
Epidural Needle	IS 14257,IS 11043:1984; IS 10654	ISO 9626; ISO 7864,	ASTM F640; ASTM F1841,ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 7864; DIN EN ISO 9626	-
Biopsy Needle	IS 14257,IS 10654	ISO 11070 (Sterile single-use intravascular introducers, dilators and guidewires; applicable principles); ISO 7864,ISO 9626	ASTM F640; ASTM F1841,ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 7864	-
Blood Collection Needle	IS 16097,IS 10654	ISO 9626; ISO 7864,	ASTM F640; ASTM F1841,ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 7864; DIN EN ISO 9626	-
Insulin Pen Needle	IS 12227,IS 10654	ISO 11608-2 (Needles for pen injectors),ISO 7864; ISO 9626	ASTM F640; ASTM F1841,ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 11608-2 (Needles for pen injectors)	-
IV Needle	IS 12153,IS 10654:2018*	ISO 7864; ISO 9626,	ASTM F640; ASTM F1841,ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 7864	-
Cannula Needle	IS 12153,IS 10654	ISO 7864,ISO 9626	ASTM F640; ASTM F1841,ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 7864	-
Dental Needle	IS 6520,IS 10654	ISO 9626; ISO 7864,	ASTM F640,ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 9626	-
Suture Needle	IS 3885,IS 10654	ISO 9626 (needle material requirements); ISO 7153-1 (Surgical instruments — metallic materials),ISO 7864	ASTM F899 (Wrought stainless steel for surgical instruments),ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 7153-1 (Surgical instruments — Metallic materials); DIN EN ISO 9626	-
Bone Marrow Aspiration Needle	IS 14257,IS 10654	ISO 7864,ISO 9626	ASTM F640; ASTM F899,ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 7864	-
Safety Needle	IS 10654,	ISO 23908; ISO 7864,ISO 9626	ASTM F640; ASTM F1841,ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 23908; DIN EN ISO 7864	AAMI TIR17 (Sharps injury prevention guidance)
Butterfly Needle	IS 16097,IS 10654	ISO 9626; ISO 11070,ISO 7864	ASTM F640; ASTM F1841,ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 9626; DIN EN ISO 7864	-
Fistula Needle	IS 10654	ISO 9626,ISO 7864	ASTM F640; ASTM F1841,ASTM F2132; ASTM F2878; ASTM F3014	DIN EN ISO 9626	-

Source field: Medical devices