Sterilization Pouches

Product / Application	BIS	ISO	ASTM	DIN	Other Standards
Self-Sealing Sterilization Pouch	IS/ISO 11607-1:2019,IS 13575	ISO 11607-1; ISO 11607-2,	ASTM F88; ASTM F1929; ASTM F1980,ASTM F1140; ASTM F1608; ASTM F2054; ASTM F2227; ASTM F2228; ASTM F3039	DIN EN 868-5 (Heat sealable pouches and reels); DIN EN ISO 11607-1	AAMI ST72; AAMI ST79
Heat Seal Sterilization Pouch	IS/ISO 11607-1:2019,IS 13575	ISO 11607-1; ISO 11607-2,	ASTM F88; ASTM F1929; ASTM F2029 (Heat sealing properties of flexible barrier materials),ASTM F1140; ASTM F1608; ASTM F2054; ASTM F2227; ASTM F2228; ASTM F3039	DIN EN 868-5; DIN EN ISO 11607-2	AAMI ST72
Flat Sterilization Pouch	IS/ISO 11607-1:2019,IS 13575	ISO 11607-1; ISO 11607-2,	ASTM F88; ASTM F1929,ASTM F1140; ASTM F1608; ASTM F2054; ASTM F2227; ASTM F2228; ASTM F3039	DIN EN 868-5	AAMI ST72
Gusseted Sterilization Pouch	IS/ISO 11607-1:2019,IS 13575	ISO 11607-1; ISO 11607-2,	ASTM F88; ASTM F2029,ASTM F1140; ASTM F1608; ASTM F2054; ASTM F2227; ASTM F2228; ASTM F3039	DIN EN 868-5	AAMI ST72
Reel Sterilization Pouch	IS/ISO 11607-1:2019,IS 13575	ISO 11607-1; ISO 11607-2,	ASTM F88; ASTM F1929,ASTM F1140; ASTM F1608; ASTM F2054; ASTM F2227; ASTM F2228; ASTM F3039	DIN EN 868-5	AAMI ST72
Tyvek Sterilization Pouch	IS/ISO 11607-1:2019,IS 13575	ISO 11607-1; ISO 11607-2,	ASTM F88; ASTM F1929; ASTM F2096,ASTM F1140; ASTM F1608; ASTM F2054; ASTM F2227; ASTM F2228; ASTM F3039	DIN EN 868-5; DIN EN ISO 11607-1	AAMI ST72
Steam Sterilization Pouch	IS/ISO 11607-1:2019,IS 13575	ISO 11607-1; ISO 11607-2; ISO 17665 (Moist heat sterilization process validation),	ASTM F88; ASTM F1980; ASTM F2029,ASTM F1140; ASTM F1608; ASTM F2054; ASTM F2227; ASTM F2228; ASTM F3039	DIN EN 868-5; DIN EN ISO 17665 (Moist heat sterilization process validation)	AAMI ST79 (Steam sterilization practices); AAMI ST72
EO Sterilization Pouch	IS/ISO 11607-1:2019,IS 13575	ISO 11607-1; ISO 11607-2; ISO 11135 (Ethylene oxide sterilization validation),	ASTM F88; ASTM F1929; ASTM F1980,ASTM F1140; ASTM F1608; ASTM F2054; ASTM F2227; ASTM F2228; ASTM F3039	DIN EN 868-5; DIN EN ISO 11135 (Ethylene oxide sterilization validation)	AAMI ST79; AAMI TIR17 (Chemical sterilization considerations where applicable)
Plasma Sterilization Pouch	IS/ISO 11607-1:2019,IS 13575	ISO 11607-1; ISO 11607-2,	ASTM F88; ASTM F1929; ASTM F1980,ASTM F1140; ASTM F1608; ASTM F2054; ASTM F2227; ASTM F2228; ASTM F3039	DIN EN 868-5; DIN EN ISO 14937 (Sterilization of healthcare products — General requirements for characterization of sterilizing agents)	AAMI ST79; AAMI ST72
Sterilization Wrap	IS/ISO 11607-1:2019; IS 17387,IS 13575	ISO 11607-1; ISO 11607-2	ASTM F1670 (Barrier material resistance to blood penetration where applicable); ASTM F1671/F1671M (Resistance to penetration by blood-borne pathogens where applicable),ASTM F1140; ASTM F1608; ASTM F2054; ASTM F2227; ASTM F2228; ASTM F3039; ASTM F88	DIN EN ISO 11607-1; DIN EN 868 series	AAMI ST79
Sterilization Reel	IS/ISO 11607-1:2019,IS 13575	ISO 11607-1; ISO 11607-2,	ASTM F88; ASTM F1929; ASTM F2029,ASTM F1140; ASTM F1608; ASTM F2054; ASTM F2227; ASTM F2228; ASTM F3039	DIN EN 868-5	AAMI ST72
Sterilization Container System	IS/ISO 11607-1:2019 and IS/ISO 11607-2:2019,IS 13575	ISO 11607-1; ISO 11607-2,	ASTM D4169; ASTM F2096; ASTM F1980,ASTM F1140; ASTM F1608; ASTM F2054; ASTM F2227; ASTM F2228; ASTM F3039; ASTM F88	DIN EN ISO 11607-1; DIN EN ISO 11607-2	AAMI ST77; AAMI ST79

Source field: Medical devices