Surgical Gloves

Product / Application	BIS	ISO	ASTM	DIN	Other Standards
Latex Surgical Gloves	IS 13422,IS 4148	ISO 10282 (Single-use sterile rubber surgical gloves — Specification),ISO 11193	ASTM D3577 (Standard Specification for Rubber Surgical Gloves),ASTM D5151; ASTM D5712; ASTM D6124; ASTM D6977; ASTM D7102; ASTM D7160; ASTM D7161; ASTM D7427	DIN EN ISO 10282 (Single-use sterile rubber surgical gloves — Specification)	AAMI/ANSI ST79 (Sterile processing relevance)
Powdered Latex Gloves	IS 13422,IS 4148	ISO 10282 (Single-use sterile rubber surgical gloves — Specification),ISO 11193	ASTM D3577; ASTM D6124 (Residual Powder on Medical Gloves),ASTM D5151; ASTM D5712; ASTM D6977; ASTM D7102; ASTM D7160; ASTM D7161; ASTM D7427	DIN EN ISO 10282; DIN EN 455-3 (Requirements and testing for biological evaluation)	-
Powder-Free Latex Gloves	IS 13422,IS 4148	ISO 10282 (Single-use sterile rubber surgical gloves — Specification),ISO 11193	ASTM D3577,ASTM D5151; ASTM D5712; ASTM D6124; ASTM D6977; ASTM D7102; ASTM D7160; ASTM D7161; ASTM D7427	DIN EN ISO 10282; DIN EN 455 series	-
Nitrile Surgical Gloves	IS 15354,IS 13422; IS 4148	ISO 10282 (Single-use sterile rubber surgical gloves — Specification); ISO 11193-1 (Medical examination gloves, where applicable),	ASTM D6319 (Standard Specification for Nitrile Examination Gloves); ASTM D6977 (Resistance of Medical Gloves to Permeation by Chemotherapy Drugs, where applicable),ASTM D3577; ASTM D5151; ASTM D5712; ASTM D6124; ASTM D7102; ASTM D7160; ASTM D7161; ASTM D7427	DIN EN ISO 10282; DIN EN 455 series (Medical gloves for single use)	-
Neoprene Surgical Gloves	IS 15354,IS 13422; IS 4148	ISO 10282 (Single-use sterile rubber surgical gloves — Specification),ISO 11193	ASTM D6977 (Chemotherapy Drug Permeation Resistance of Medical Gloves); ASTM D3577 principles,ASTM D5151; ASTM D5712; ASTM D6124; ASTM D7102; ASTM D7160; ASTM D7161; ASTM D7427	DIN EN ISO 10282	-
Polyisoprene Surgical Gloves	IS 15354,IS 13422; IS 4148	ISO 10282 (Single-use sterile rubber surgical gloves — Specification),ISO 11193	ASTM D3577; ASTM D6977 where chemical resistance is required,ASTM D5151; ASTM D5712; ASTM D6124; ASTM D7102; ASTM D7160; ASTM D7161; ASTM D7427	DIN EN ISO 10282	-
Sterile Surgical Gloves	IS 13422,IS 4148	ISO 10282 (Single-use sterile rubber surgical gloves — Specification),ISO 11193	ASTM D3577; ASTM D5151 (Detection of Holes in Medical Gloves); ASTM F88 (Package Seal Strength),ASTM D5712; ASTM D6124; ASTM D6977; ASTM D7102; ASTM D7160; ASTM D7161; ASTM D7427	DIN EN ISO 10282; DIN EN 455-1 (Freedom from holes); DIN EN 455-2 (Physical properties)	AAMI ST79 (Sterilization processing environment)
Orthopaedic Surgical Gloves	IS 13422,IS 4148	ISO 10282 (Single-use sterile rubber surgical gloves — Specification),ISO 11193	ASTM D3577; ASTM D5151,ASTM D5712; ASTM D6124; ASTM D6977; ASTM D7102; ASTM D7160; ASTM D7161; ASTM D7427	DIN EN ISO 10282	-
Double Layer Surgical Gloves	IS 13422,IS 4148	ISO 10282 (Single-use sterile rubber surgical gloves — Specification),ISO 11193	ASTM D3577; ASTM D5151,ASTM D5712; ASTM D6124; ASTM D6977; ASTM D7102; ASTM D7160; ASTM D7161; ASTM D7427	DIN EN ISO 10282	-
Radiation Protection Gloves	IS 13422; IS 4148	ISO 10282 (Glove material requirements where applicable); ISO 21420 (Protective gloves — General requirements); ISO 61331-3	ASTM F2547 (Protective Gloves Against Ionizing Radiation); ASTM D1051 (Rubber Insulation — where applicable),ASTM D3577; ASTM D5151;	DIN EN 61331-3 (Protective devices against diagnostic medical X-radiation — Protective clothing and gloves); DIN EN ISO 21420	IEC 61331 (Protective devices against diagnostic medical X-radiation)

Product / Application	BIS	ISO	ASTM	DIN	Other Standards
		(Protective devices against diagnostic medical X-radiation — Flexible protective materials),ISO 11193	ASTM D5712; ASTM D6124; ASTM D6977; ASTM D7102; ASTM D7160; ASTM D7161; ASTM D7427	(Protective gloves — General requirements)	

Source field: Medical devices