Syringes

Product / Application	BIS	ISO	ASTM	DIN	Other Standards
Disposable Hypodermic Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 7886-1 (Sterile hypodermic syringes for single use — Part 1: Syringes for manual use),ISO 8537	ASTM F640 (Radiopacity testing of medical devices where applicable); ASTM D4169 (shipping validation where applicable); ASTM F88 (seal strength for packaged sterile syringes),ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 7886-1 (Sterile hypodermic syringes for single use — Part 1: Syringes for manual use)	-
Sterile Disposable Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 7886-1; ISO 11607-1 (Packaging for terminally sterilized medical devices),ISO 8537	ASTM F1980 (Accelerated aging of sterile barrier systems); ASTM F88 (seal strength); ASTM F1929 (package leak detection),ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 7886-1; DIN EN ISO 11607-1 (Packaging for terminally sterilized medical devices)	AAMI ST72 (Sterile packaging validation where applicable)
Insulin Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 8537 (Sterile syringes for insulin, single use),ISO 7886	ASTM F640 (Radiopacity where applicable); ASTM F88; ASTM F1980,ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 8537 (Sterile syringes for insulin, single use)	AAMI/ANSI ST72 (Sterile packaging where applicable)
Tuberculin Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 7886-1 (Sterile hypodermic syringes for single use),ISO 8537	ASTM F88; ASTM F1980,ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 7886-1	-
Oral Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 7886-1; ISO 80369-3 (Small-bore connectors for enteral applications),ISO 8537	ASTM F88; ASTM F1980,ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 7886-1; DIN EN ISO 80369-3 (Enteral small-bore connectors)	-
Enteral Feeding Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 80369-3 (Enteral small-bore connectors); ISO 7886-1,ISO 8537	ASTM F88; ASTM F1980; ASTM D4169 (transport testing where applicable),ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 80369-3; DIN EN ISO 7886-1	AAMI/ANSI CN27 (Enteral connectors where applicable)
Catheter Tip Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 7886-1,ISO 8537	ASTM F88; ASTM F1980,ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 7886-1	-
Luer Lock Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 80369-7 (Small-bore connectors for intravascular/hypodermic applications); ISO 7886-1,ISO 8537	ASTM F640; ASTM F88,ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 80369-7 (Small-bore connectors for intravascular/hypodermic applications); DIN EN ISO 7886-1	AAMI/ANSI ISO 80369 adoption (small-bore connectors)
Luer Slip Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 80369-7; ISO 7886-1,ISO 8537	ASTM F640; ASTM F88,ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 80369-7; DIN EN ISO 7886-1	AAMI/ANSI ISO 80369 adoption
Prefilled Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS	ISO 11040-4 (Glass barrels for injectables); ISO 11040-6 (Plastic syringes),ISO 7886; ISO 8537	ASTM F1980 (stability/aging); ASTM F88 (container closure integrity); ASTM F2096 (gross leak testing),ASTM D4169; ASTM D7386;	DIN EN ISO 11040-4 (Glass barrels for injectables); DIN EN ISO 11040-6 (Plastic barrels)	AAMI TIR57 (Principles for combination products where applicable)

Product / Application	BIS	ISO	ASTM	DIN	Other Standards
	11252		ASTM F2503; ASTM F3014		
Safety Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 7886-1; ISO 23908 (Sharps injury protection — safety-engineered devices),ISO 8537	ASTM F640; ASTM F88; ASTM F1980,ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 7886-1; DIN EN ISO 23908 (Sharps injury protection devices)	AAMI/ANSI ST72; AAMI TIR17 (Sharps safety guidance)
Auto Disable (AD) Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 7886-3 (Auto-disable syringes for fixed-dose immunization),ISO 8537	ASTM F640; ASTM F1980,ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 7886-3 (Auto-disable syringes for fixed-dose immunization)	AAMI/ANSI ST72
Irrigation Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 7886-1,ISO 8537	ASTM F88; ASTM F1980,ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 7886-1	-
Dental Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 7886-1; ISO 9997 (Dental cartridge syringes),ISO 8537	ASTM F88; ASTM F1980,ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 7886-1; DIN EN ISO 9997 (Dental cartridge syringes)	-
Glass Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 11040-4 (Glass barrels for injectables),ISO 7886; ISO 8537	ASTM E438 (Glass in laboratory apparatus); ASTM C149 (glass chemical durability where applicable),ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 11040-4	-
Reusable Syringe	IS 10258 (Sterile Single-Use Hypodermic Syringes); IS 11252	ISO 7886-2 (Reusable syringes for manual use),ISO 8537	ASTM F88; ASTM F1980,ASTM D4169; ASTM D7386; ASTM F2503; ASTM F3014	DIN EN ISO 7886-2 (Reusable syringes for manual use)	AAMI ST79 (Reusable medical device processing where applicable)

Source field: Medical devices