Rubber Gloves

Product / Application	BIS (IS)	ISO	ASTM	DIN	Other Standards
Surgical Latex Gloves	IS 13422:2024; IS 1741; IS 3565; IS 4148; IS 5190; IS 3708; IS 3400	ISO 10282 (Single-use Sterile Rubber Surgical Gloves — Specification); ISO 11193-1 (Single-use Medical Examination Gloves — Specification for Rubber Latex Gloves),ISO 2004; ISO 7439	ASTM D3577 (Rubber Surgical Gloves); ASTM D5151 (Detection of Holes in Medical Gloves); ASTM D6124 (Residual Powder on Medical Gloves); ASTM D6319 (Nitrile Examination Gloves, where applicable); ASTM D1076; ASTM D3578; ASTM D3492; ASTM D5250; ASTM D4679	DIN EN 455-1 (Medical Gloves — Freedom from Holes); DIN EN 455-2 (Physical Properties); DIN EN 455-3 (Biological Evaluation); DIN EN ISO 374-5 (Protection Against Microorganisms)	FDA 21 CFR 878 (Medical Gloves); USP Class VI (Biocompatibility); EN/European Medical Device Requirements (where applicable)
Examination Latex Gloves	IS 15354; IS 15354 (Part 1); IS 15354 (Part 2); IS 1741; IS 3565; IS 4148; IS 13422; IS 5190; IS 3708; IS 3400	ISO 11193-1 (Single-use Medical Examination Gloves — Specification for Rubber Latex Gloves); ISO 23529 (Rubber — General Procedures for Preparing and Conditioning Test Pieces),ISO 2004; ISO 10282; ISO 7439	ASTM D3578 (Rubber Examination Gloves); ASTM D5151 (Detection of Holes in Medical Gloves); ASTM D6124 (Residual Powder on Medical Gloves); ASTM D412 (Tensile Properties of Rubber); ASTM D1076; ASTM D3577; ASTM D3492; ASTM D6319; ASTM D5250; ASTM D4679	DIN EN 455-1 (Medical Gloves — Freedom from Holes); DIN EN 455-2 (Physical Requirements); DIN EN 455-3 (Biological Evaluation); DIN EN ISO 374-5 (Microorganism Protection)	FDA 21 CFR 800 Series (Medical Gloves); USP Class VI; CDC Healthcare Glove Guidelines
Nitrile Examination Gloves	IS 15354; IS 15354 (Part 1); IS 15354 (Part 2); IS 1741; IS 3565; IS 4148; IS 13422; IS 5190; IS 3708; IS 3400	ISO 11193-1 (Single-use Medical Examination Gloves); ISO 11193-2 (Gloves Made from Nitrile Rubber),ISO 2004; ISO 10282; ISO 7439	ASTM D6319 (Nitrile Examination Gloves for Medical Applications); ASTM D5151 (Detection of Holes in Medical Gloves); ASTM D6124 (Residual Powder on Medical Gloves); ASTM D412 (Tensile Properties of Rubber); ASTM D1076; ASTM D3577; ASTM D3578; ASTM D3492; ASTM D5250; ASTM D4679	DIN EN 455-1 (Medical Gloves — Freedom from Holes); DIN EN 455-2 (Physical Properties); DIN EN ISO 374-1 (Chemical Protective Gloves); DIN EN ISO 374-5 (Protection Against Microorganisms)	FDA 21 CFR 880.6250 (Patient Examination Gloves); USP Class VI; CDC Healthcare PPE Guidelines
Neoprene Examination Gloves	IS 15354; IS 1741; IS 3565; IS 4148; IS 13422; IS 5190; IS 3708; IS 3400	ISO 11193-1 (Single-use Medical Examination Gloves),ISO 2004; ISO 10282; ISO 7439	ASTM D6977 (Cut Resistance of Medical Gloves, where applicable); ASTM D5151 (Detection of Holes in Medical Gloves); ASTM D412 (Tensile Properties of Rubber); ASTM D471 (Effect of Liquids on Rubber); ASTM D1076; ASTM D3577; ASTM D3578; ASTM D3492; ASTM D6319; ASTM D5250; ASTM D4679	DIN EN 455-1 (Medical Gloves — Freedom from Holes); DIN EN 455-2 (Physical Properties); DIN EN ISO 374-1 (Chemical Resistance); DIN EN ISO 374-5 (Microorganism Resistance)	FDA Medical Glove Requirements; USP Class VI; CDC Healthcare PPE Guidelines
Vinyl Examination Gloves	IS 15354; IS 1741; IS 3565; IS 4148; IS 13422; IS 5190; IS 3708; IS 3400	ISO 11193-1 (Single-use Medical Examination Gloves),ISO 2004; ISO 10282; ISO 7439	ASTM D5250 (Polyvinyl Chloride Gloves for Medical Applications); ASTM D5151 (Detection of Holes in Medical Gloves); ASTM D6124 (Residual Powder on Medical Gloves); ASTM D412 (Tensile Properties of Rubber and Thermoplastic Elastomers); ASTM D1076; ASTM D3577; ASTM D3578; ASTM D3492; ASTM D6319; ASTM D4679	DIN EN 455-1 (Medical Gloves — Freedom from Holes); DIN EN 455-2 (Physical Properties); DIN EN ISO 374-5 (Microorganism Resistance); DIN EN ISO 527-1 (Tensile Properties of Polymer Materials)	FDA 21 CFR 880.6250; USP Class VI; CDC Healthcare PPE Guidelines
Industrial Chemical Gloves	IS 15331; IS 1741; IS 3565; IS 4148; IS 13422; IS 5190; IS 3708; IS 3400	ISO 374-1 (Protective Gloves Against Dangerous Chemicals and Micro-organisms); ISO 374-4 (Resistance to Degradation by Chemicals); ISO 374-5 (Protection Against Micro-organisms),ISO 2004; ISO 10282; ISO 11193-1; ISO	ASTM F739 (Permeation Resistance of Protective Clothing Materials to Chemicals); ASTM F1383 (Permeation Resistance of Protective Clothing Materials to Liquid Chemicals Under Conditions of Intermittent Contact); ASTM D6978 (Resistance of Medical	DIN EN ISO 374-1 (Protective Gloves Against Dangerous Chemicals); DIN EN ISO 374-5 (Protection Against Microorganisms); DIN EN 388 (Mechanical Risks); DIN EN ISO 188 (Ageing Resistance, where rubber applies)	OSHA PPE Requirements; NFPA 1991 (Chemical Protective Ensembles, where applicable); ANSI/ISEA 105 (Hand Protection Classification)

Product / Application	BIS (IS)	ISO	ASTM	DIN	Other Standards
		7439	Gloves to Permeation by Chemotherapy Drugs, where applicable); ASTM D412 (Tensile Properties of Rubber); ASTM D1076; ASTM D3577; ASTM D3578; ASTM D3492; ASTM D6319; ASTM D5250; ASTM D4679		
Electrical Insulating Gloves	IS 13774:2021; IS 4770:1991; IS 1741; IS 3565; IS 4148; IS 13422; IS 5190; IS 3708; IS 3400	,ISO 2004; ISO 10282; ISO 11193-1; ISO 7439	ASTM D120 (Rubber Insulating Gloves); ASTM F496 (In-Service Care of Insulating Gloves and Sleeves); ASTM F1236 (Visual Inspection of Electrical Protective Rubber Products); ASTM D149 (Dielectric Breakdown Voltage and Dielectric Strength of Solid Electrical Insulating Materials); ASTM D1076; ASTM D3577; ASTM D3578; ASTM D3492; ASTM D6319; ASTM D5250; ASTM D4679	DIN EN IEC 60903 (Live Working — Electrical Insulating Gloves); DIN EN 50237 (Electrical Insulating Gloves with Mechanical Protection, where applicable); DIN EN 388 (Mechanical Protection); DIN EN ISO 374-5 (Chemical/Microbial Protection, where applicable)	OSHA 29 CFR 1910.137 (Electrical Protective Equipment); NFPA 70E (Electrical Safety in Workplace)
Household Rubber Gloves	IS 4148 (Rubber Household Gloves); IS 1741; IS 3565; IS 13422; IS 5190; IS 3708; IS 3400	ISO 11193-1 (Single-use Rubber Gloves, where applicable),ISO 2004; ISO 10282; ISO 7439	ASTM D4679 (Rubber General-Purpose, Household or Industrial Gloves); ASTM D412 (Tensile Properties of Rubber); ASTM D2240 (Durometer Hardness of Rubber); ASTM D471 (Effect of Liquids on Rubber); ASTM D1076; ASTM D3577; ASTM D3578; ASTM D3492; ASTM D6319; ASTM D5250	DIN EN ISO 374-1 (Chemical Protective Gloves, where applicable); DIN EN ISO 374-5 (Microorganism Protection); DIN EN 388 (Mechanical Protection); DIN ISO 37 (Tensile Properties)	FDA 21 CFR 177 (Food Handling Applications, where applicable); ANSI/ISEA 105 (Hand Protection Classification); OSHA PPE Requirements
Food Handling Gloves	IS 4148 (Rubber Gloves); IS 10141 (Food Handling Applications, where applicable); IS 1741; IS 3565; IS 13422; IS 5190; IS 3708; IS 3400	ISO 22000 (Food Safety Management Systems); ISO 21420 (Protective Gloves — General Requirements and Test Methods),ISO 2004; ISO 10282; ISO 11193-1; ISO 7439	ASTM D5250 (PVC Gloves for Medical Applications, where applicable); ASTM D6319 (Nitrile Gloves, where applicable); ASTM D5151 (Detection of Holes in Medical Gloves); ASTM D412 (Tensile Properties of Rubber); ASTM D1076; ASTM D3577; ASTM D3578; ASTM D3492; ASTM D4679	DIN EN 1186 (Materials and Articles in Contact with Foodstuffs); DIN EN 455-2 (Physical Properties, where medical grade applies); DIN EN ISO 374-5 (Microorganism Protection); DIN ISO 37 (Tensile Properties)	FDA 21 CFR 177; NSF/ANSI 51; HACCP Food Safety Requirements

Source field: Rubber and Rubber products