Medical Textiles

Product / Application	BIS (IS)	ISO	ASTM	DIN	Other Standards
Surgical Gown; Surgical Drape	IS 17334 (Medical Protective Clothing); IS 17423 (Protective Clothing for Healthcare Applications),IS 16252; IS 15570; IS 6359	ISO 13795 (Surgical Drapes, Gowns and Clean Air Suits Used as Medical Devices for Patients, Clinical Staff and Equipment),ISO 10993; ISO 11135; ISO 13485; ISO 22610; ISO 22612	ASTM F2407 (Surgical Gowns Intended for Use in Healthcare Facilities); ASTM F1670 (Resistance to Penetration by Synthetic Blood); ASTM F1671 (Resistance Against Blood-Borne Pathogens); ASTM D5034 (Breaking Strength and Elongation),ASTM F2100; ASTM F1868; ASTM F3050; ASTM F2299	DIN EN 13795-1 (Surgical Clothing and Drapes — Requirements and Test Methods); DIN EN ISO 13934-1 (Tensile Properties of Fabrics); DIN EN ISO 811 (Resistance to Water Penetration); DIN EN ISO 22610 (Resistance to Wet Bacterial Penetration)	FDA 21 CFR 878 (Surgical Apparel, where applicable); AAMI PB70 (Liquid Barrier Performance of Protective Apparel); CDC Healthcare PPE Guidelines
Surgical Mask	IS 16289,IS 16252; IS 17334; IS 17423; IS 15570; IS 6359	ISO 22609 (Medical Face Masks — Test Method for Resistance Against Penetration by Synthetic Blood); ISO 11737-1 (Sterilization of Health Care Products — Microbiological Methods),ISO 10993; ISO 11135; ISO 13485; ISO 22610; ISO 22612	ASTM F2100 (Performance of Materials Used in Medical Face Masks); ASTM F2101 (Bacterial Filtration Efficiency of Medical Face Mask Materials); ASTM F1862 (Resistance of Medical Face Masks to Penetration by Synthetic Blood); ASTM D737 (Air Permeability),ASTM F1671; ASTM F1868; ASTM F2407; ASTM F3050; ASTM F2299	DIN EN 14683 (Medical Face Masks — Requirements and Test Methods); DIN EN ISO 9237 (Air Permeability of Textiles); DIN EN ISO 22612 (Resistance to Microbial Penetration, where applicable); DIN EN ISO 10993-5 (Biological Evaluation, where skin contact applies)	FDA 21 CFR 878 (Surgical Masks); NIOSH Requirements (Respiratory Protection Applications); AAMI PB70 (Barrier Performance, where applicable)
N95 Respirator	IS 9473; IS 17423 (where applicable),IS 16252; IS 17334; IS 15570; IS 6359	ISO 16900-1 (Respiratory Protective Devices — Methods of Test and Test Equipment); ISO 17420-1 (Respiratory Protective Devices — Performance Requirements),ISO 10993; ISO 11135; ISO 13485; ISO 22610; ISO 22612	ASTM F2299 (Aerosol Filtration Efficiency of Nonwoven Filter Media); ASTM F2100 (Medical Face Mask Materials, where applicable); ASTM F1862 (Resistance to Synthetic Blood, where applicable); ASTM D737 (Air Permeability),ASTM F1671; ASTM F1868; ASTM F2407; ASTM F3050	DIN EN 149 (Respiratory Protective Devices — Filtering Half Masks); DIN EN ISO 16900 (Respiratory Protective Devices — Test Methods); DIN EN ISO 9237 (Air Permeability); DIN EN ISO 16890 (Air Filter Testing, where applicable)	NIOSH 42 CFR Part 84 (Respiratory Protective Devices); CDC Respiratory Protection Guidelines; OSHA Respiratory Protection Requirements
Sterilization Wrap	IS 17387 (Packaging Materials for Terminally Sterilized Medical Devices),IS 16252; IS 17334; IS 17423; IS 15570; IS 6359	ISO 11607-1 (Packaging for Terminally Sterilized Medical Devices — Requirements for Materials, Sterile Barrier Systems and Packaging Systems),ISO 10993; ISO 11135; ISO 13485; ISO 22610; ISO 22612	ASTM F1671 (Resistance of Materials Used in Protective Clothing Against Blood-Borne Pathogens); ASTM F88 (Seal Strength of Flexible Barrier Materials, where pouch/wrap sealing applies); ASTM F1929 (Detecting Seal Leaks in Porous Medical Packaging by Dye Penetration); ASTM D882 (Tensile Properties of Thin Plastic Sheeting),ASTM F2100; ASTM F1868; ASTM F2407; ASTM F3050; ASTM F2299	DIN EN ISO 11607-1 (Packaging for Terminally Sterilized Medical Devices); DIN EN ISO 11607-2 (Validation of Forming, Sealing and Assembly Processes); DIN EN 868-5 (Packaging Materials and Systems for Sterilization); DIN EN ISO 16775 (Guidance on ISO 11607)	FDA Medical Device Packaging Requirements; USP Packaging Requirements; AAMI ST79 (Sterilization and Sterility Assurance Practices)
Medical Gauze	IS 758,IS 16252; IS 17334; IS 17423; IS 15570; IS 6359	ISO 13485 (Medical Devices — Quality Management Systems); ISO 10993-1 (Biological Evaluation of Medical Devices, where applicable),ISO 10993; ISO 11135; ISO 22610; ISO 22612	ASTM D5034 (Breaking Strength and Elongation of Textile Fabrics); ASTM D3776 (Mass Per Unit Area of Fabric); ASTM D5035 (Breaking Force and Elongation of Textile Fabrics); ASTM E2149 (Antimicrobial Activity of Immobilized Antimicrobial Agents, where applicable),ASTM F1671; ASTM F2100; ASTM F1868; ASTM F2407; ASTM F3050; ASTM F2299	DIN EN 14079 (Non-active Medical Devices — Requirements for Cotton Gauze); DIN EN ISO 10993-5 (Biological Evaluation — Cytotoxicity); DIN EN ISO 9073-3 (Tensile Strength of Nonwovens); DIN EN ISO 9073-6 (Absorption Capacity)	USP Requirements for Medical Gauze; FDA Medical Device Requirements; AAMI Guidelines for Surgical Materials

Source field: Textile and Textile Products