

Medical Device Recall Report

Period: 2026_Week2 (2026-07-08 to 2026-07-14)

Summary

Metric	Count
Total FDA Recalls	2
Total ECRI Alerts	1
Hospital Matches	3
Possible Matches	0
No Match	0

Match Detail

Source	Recall #	Asset #	Manufacturer	Model	Score	Status	Method
FDA	Z-1235-2026	A-1002	Zenith Health	VT-750	100%	Matched	UDI
FDA	Z-1234-2026	A-1001	Acme Medical	IP-3000	100%	Matched	UDI
ECRI	ECRI-8891	A-1007	Zenith Health	US-900	98%	Matched	Model Number

FDA Recalls

Recall #	Class	Product	Manufacturer	Reason	Date Initiated	Status
Z-1235-2026	Class II	Ventilator VT-750	Zenith Health	Battery indicator may show incorrect charge level.	2026-07-11	Ongoing
Z-1234-2026	Class I	Infusion Pump IP-3000	Acme Medical	Pump may stop infusion without alarming.	2026-07-09	Ongoing

ECRI Alerts

Alert #	Title	Manufacturer	Model	Hazard	Date Issued	Status
ECRI-8891	Ultrasound Machine US-900 probe connector failure risk	Zenith Health	US-900	High	2026-07-10	Active