

# CM and DM EMI Analysis

30 September, 2014

Manoj Gulati



# Why phase information is lost?

- Phase information is lost in spectrum analyzers as they don't preserve phase information while going through heterodyning process. [Agilent Report on Spectrum Analyzers]. They are meant to measure the power of the spectrum and shows only magnitude vs. frequency with in full range of instrument.
- Vector Signal Analyzer measure the magnitude and phase of an input signal at a single frequency with in the IF bandwidth. Primary use is to make in channel measurements like error vector magnitude, spectral flatness on known signals.
- Latest signal Analyzers provides the function of both of them. (Depends on model).

NB: Need to check this with Agilent if our N9000CXA support this.



# Types of Spectrum Analyzers

## FFT Analyzer



# Swept Analyzer

Working principal of majority of spectrum analyzers.



Idea: USRP as a Data Acquisition System to Capture CM and DM mode Conducted EMI

- Currently we are capturing time domain data using USRP and doing FFT offline on MATLAB. So we have phase information in that data.
- But we were capturing differential mode EMI only. We can modify HPF to work in split phase and then it can be used to capture Vphase and Vneutral.
- Numerical algorithm in MATLAB can separate CM and DM components.

# References

- Agilent Guide : Fundamentals of Spectrum Analysers
- Agilent Video Tutorials : Fundamentals of Spectrum Analyzers