

Equipment Safety & Facility Compliance Audit

Location / Facility Identifier:

Building-4B / Cleanroom 2

Safety Classification Level:

Level 1 - Low Risk

Inspection Verdict (Radio):

☒ Passed ☐ Requires Follow-up ☐ Critical Fail

Emergency Stops Operational:

☐

Ventilation Airflow Passed:

☐

Corrective Actions Required:

Calibrate primary differential pressure sensor before Friday.