Certificate of Calibration

CALIBRATION
NVLAP LAB CODE 600214-0

Model:	ADT685-05-CP1K-PSI-N	**Report Number:**	N26062249
Description:	Digital Pressure Gauge	**Calibration Date:**	6/22/2026
Serial Number:	030250050090	**Due Date:**	6/21/2027
Pressure Range:	(-15 to 1000)psi;(60 to 110)kPa.a	**Temperature:**	22.4°C
Accuracy(%FS):	0.05%FS	**Humidity:**	52%RH
Status:	New	**Certificate Date:**	6/22/2026
Procedure:	ADT685;Lab3.4.2;ADT685V2.0.6.0	**Data Type:**	As-Left
		Data Results:	In Tolerance

Customer:

Location of Calibration:
Additel Corporation
2900 Saturn Street #B
Brea CA 92821
United States

Reference Standards used in this calibration are traceable to the SI Units through National Institute of Standards and Technology (NIST) or other recognized National Metrology Institutes (NMI's). This calibration is compliant to ISO/IEC 17025:2017 and ANSI/NCSL Z540-1:R2002.

By default, Additel reports the measurement result and its associated measurement uncertainty. When the measurement result is outside Additel's published specifications, it is indicated by a fail condition without taking the associated measurement uncertainty into consideration. Alternatively, if a compliance decision is desired taking the measurement uncertainty into account, Additel utilizes the ANSI Z540.3 Handbook Method 6 to make statements of conformity to specifications.

The results on this calibration report apply only to the item (s) calibrated or tested. Uncertainties of the measurements are based upon a coverage factor of k=2 providing an approximate 95% confidence level. The recommended calibration due date is included in the report only if it is specified by the customer.

This report shall not be used by the customer to claim product certification, approval, or endorsement by NVLAP, NIST, or any agency of the U.S. government. This report shall not be reproduced, except in full without written approval of Additel Corporation.

Standards Used

Model	Description	Serial Number	Cal Date	Due Date
ADT160A-AP120R-KPA-N	Additel Quartz Pressure Module	CDPR201C0002	6/15/2026	10/20/2026
ADT770-100	Additel Pressure Controller	82118020004	5/7/2026	5/8/2027
ADT160A-AP1KR-PSI-N	Additel Quartz Pressure Module	CDPR18040009	3/13/2026	7/26/2026
ADT780-3K	Additel Pressure Controller	822170C0001	5/7/2026	5/7/2027

Comments:

Performed By: Kyle Bague

Approved By: Kyle Bague

Test Results

Barometric Pressure Sensor

Range:(60 to 110)kPa.a		Accuracy:±55Pa.a			S/N:BMP251202152246	
TestPoint (kPa.a)	Reference (kPa.a)	DUT (kPa.a)	Difference (kPa.a)	Specification (kPa.a)	Uncertainty (kPa.a)	Condition
60.000	60.000	59.997	-0.003	±0.055	6.1E-03	pass
80.000	80.000	79.996	-0.004	±0.055	8.1E-03	pass
100.000	100.000	99.992	-0.008	±0.055	1.0E-02	pass
110.000	110.000	109.997	-0.003	±0.055	1.1E-02	pass

High Pressure

Range:(-15 to 1000)psi		Accuracy:0.05%FS			S/N:DPSE026D320028	
TestPoint (psi)	Reference (psi)	DUT (psi)	Difference (psi)	Specification (psi)	Uncertainty (psi)	Condition
-13.00	-13.00	-13.00	0.00	±0.51	2.5E-02	pass
-7.25	-7.25	-7.25	0.00	±0.51	2.5E-02	pass
0.00	0.00	0.00	0.00	±0.51	2.5E-02	pass
250.00	250.00	249.99	-0.01	±0.51	2.5E-02	pass
500.00	500.00	500.00	0.00	±0.51	4.2E-02	pass
750.00	750.00	749.99	-0.01	±0.51	6.1E-02	pass
1000.00	1000.00	999.99	-0.01	±0.51	8.1E-02	pass
750.00	750.00	749.99	-0.01	±0.51	6.1E-02	pass
500.00	500.00	500.00	0.00	±0.51	4.2E-02	pass
250.00	250.00	249.99	-0.01	±0.51	2.5E-02	pass
0.00	0.00	0.00	0.00	±0.51	2.5E-02	pass