CERTIFICATE OF CALIBRATION

TEKTON.

Device	1/2 Inch Drive Micrometer Torque Wrench
Item#	TRQ52402
Serial #	H5 5240202447
Torque Units	ft-lb
Torque Range	40 - 200

Summary of Accuracy at Calibration

	40 ft-lb	120 ft-lb	200 ft-lb
CW	-0.3%	1.4%	2.3%
CCW	-1.3%	0.6%	2.1%

According to ISO 6789-1:2017, the maximum allowable deviation is +/- 4%.

Date of Calibration

2025-10-03

To maintain original accuracy, recalibrate every 12 months or 5000 cycles.

Set Torque	40 ft-lb		120 ft-lb		200 ft-lb	
Clockwise (CW)	Measured Values	Deviation	Measured Values	Deviation	Measured Values	Deviation
1	40.2	0.5%	122.5	2.1%	205.7	2.9%
2	40.1	0.3%	121.6	1.3%	204.8	2.4%
3	39.8	-0.5%	121.4	1.2%	204.6	2.3%
4	39.6	-1.0%	121.4	1.2%	204.3	2.2%
5	39.6	-1.0%	121.4	1.2%	203.4	1.7%
Average	39.9	-0.3%	121.7	1.4%	204.6	2.3%
Counterclockwise (CCW)	Measured Values	Deviation	Measured Values	Deviation	Measured Values	Deviation
1	39.5	-1.3%	120.8	0.7%	204.0	2.0%
2	39.5	-1.3%	120.7	0.6%	204.2	2.1%
3	39.5	-1.3%	120.6	0.5%	204.2	2.1%
4	39.5	-1.3%	120.7	0.6%	204.4	2.2%
5	39.5	-1.3%	120.6	0.5%	204.2	2.1%
Average	39.5	-1.3%	120.7	0.6%	204.2	2.1%

Test equipment

Model: PB-6036
Serial #: 047-15031901
Accuracy: +/- 1%

Test conditions

Inspector: Amanda
Temperature: 27°C
Humidity: 55%

Test standard

ISO 6789-1:2017

Scanned with CamScanner