OPNAVINST 5100.29

PERIODIC AUTOMATED EXTERNAL DEFIBULLATOR (AED) RECORD FOR YEAR _____

Installation: _____ AED Location: _____

Date of Inspection	AED Case Intact	Battery Charged "Ready for Use"	All Equipment is in AED Cabinet	Inspector Initials
01/				
02/				
03/				
04/				
05/				
06/				
07/				
08/				
09/				
10/				
11/				
12/				

OPNAV 5522/2 (JUL 2011)

Equipment Inventory:

- 1 Simplified CPR/AED instruction
- 4 Pairs of non-latex gloves in varying sizes (S, M, L, XL)
- 1 CPR facemask with some type of barrier device
- 1 Disposable razor
- 1 Pair blunt end scissors
- 1 Set adult AED pads
- 1 Set child AED pads (not necessary in areas where pediatrics are restricted)
- 1 Biohazard bag
- 1 Absorbent towel