

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

C-Star Calibration and Repairs

Date 1/7/2022 Customer University of Alaska - Seward

S/N# CST-1609DR Work order 1005514813

Standard Service

- Cleaned meter
- Replaced all seals
- Focused and aligned optics
- Performed noise test: 1 sample/sec for 60 sec
- Performed stability test: 1 sample/min for 12 hrs
- Performed temperature test
- Performed water calibration
- Shake-tested unit
- Pressure-tested unit
- Updated unit's calibration sheet
- Check for a zinc anode