

R6E8A 24-Hour Pressure Report

Latest processed 24-hour window
Pressure first · EHZ
separate · outside guides only

LATEST DUAL-CHANNEL CUTOFF August 4, 2026, 7:02 AM Eastern Time Latest processed point available from both channels.	HDF PRESSURE LEADS approximately 0.246 Pa RMS* Nominal one-second estimate; not a handbook grade.	PRESSURE EVENT Aug 4 - 6:48:57 AM EST 2.3549 Hz · 1.0 sec detector window	EXTERNAL GUIDE RESULT Not yet gradeable No percentage or tier until the same required metric is measured.
--	---	---	---

Plain-language bottom line

Pressure activity is present in the HDF record. The current feed supports a nominal pressure estimate, but not a truthful percent-above, danger label, pass/fail result or monitoring tier under the four published guides.

Four published-guide comparison board

Each row keeps the guide’s own units and duration. No same-station or peer-station baseline is used.

PUBLISHED GUIDE	OUTSIDE REFERENCE	R6E8A RESULT	WHY
1 · ISO 7196 Infrasound pressure	G-weighted pressure for 1-20 Hz. It is a measurement method and sets no danger limit.	N/A · dB(G) not yet computed	Needs station-response correction to Pa and validated G weighting.
2 · ANSI/ASA + ASHRAE Hotel room	Hotel room/suite target NC/RC 30; approximately 35 dBA and 60 dBC. Project target may vary ±5.	N/A · NC/RC not measured	Needs a calibrated room survey with the required octave bands and positions.
3 · Defra NANR45 Low frequency	Indoor five-minute 1/3-octave night curve: 92 to 34 dB across 10-160 Hz.	N/A · full band set unavailable	R6E8A cannot supply the required 50-160 Hz bands or full indoor procedure.
4 · WHO + ISO 1996 Night environment	30 dBA bedroom overnight; 40 dBA annual outside guideline; 55 dBA annual interim target.	N/A · dBA averages not measured	Needs calibrated A-weighted sound and each guide’s location and averaging period.

R6E8A event record

HDF pressure leads
EHZ motion remains
separate

HDF · AIR PRESSURE		EHZ · VERTICAL MOTION	
Availability	Available	Availability	Available
Highest detector event	Aug 4 - 6:48:57 AM EST	Highest detector event	Aug 4 - 12:51:51 AM EST
Nominal pressure	approximately 0.246 Pa RMS*	Instrument reading	20,107.2 counts
Dominant frequency	2.3549 Hz	Dominant frequency	5.0426 Hz
Detector window	1.0 seconds	Detector window	8.0 seconds
Outside-guide result	N/A · unmatched metric	Acoustic guides	Not applicable to EHZ

Preserved recent HDF pressure events

DATE	HDF EVENT	NOMINAL PA RMS*	FREQUENCY	DETECTOR WINDOW	OUTSIDE GUIDE
2026-08-04	Aug 4 - 6:48:57 AM EST	approximately 0.246 Pa RMS*	2.3549 Hz	1.0 sec	N/A

Monitoring-priority framework

TIER 1	TIER 2	TIER 3	TIER 4
Within named guide	Intermittent exceedance	Prolonged exceedance	Extended exceedance
Matched result at/below guide	Each interval under 30 min	30-119 continuous min	120+ continuous min

Current tier: N/A. A detector window is not an outside-guide exceedance interval.

Reading controls

HDF is air pressure; EHZ is vertical motion; their amplitudes are never combined. Detector windows locate events, not time above a guide. Missing samples remain N/A.

*Nominal Pa = one-second HDF RMS counts ÷ 56,000 (manufacturer estimate, ±10%); response-uncorrected and unweighted; not a guide or safety grade.

*Account for up to 30 minutes of lag. For personal verification, use the Raspberry Shake source service. This report documents instrument records and published-guide readiness; it does not identify a source, cause, person, medical effect or safety condition.