Table of Contents

iMotions Lab > Analysis > Signal Processing (R Notebooks) > Facial Expression Analysis > AFFDEX

Facial Expression Analysis: Why does AFFDEX show low QA in subject replay? 2

Confidential © 2023 Redistribution is not permitted without written permission from iMotions

Facial Expression Analysis: Why does AFFDEX show low QA in subject replay?

This article reviews why AFFDEX shows low Quality for subject replays after data collection.

The sampling rate for Affectiva AFFDEX depends on the camera you are using to collect data. For example, if you are using the Logitech C920, iMotions assumes that AFFDEX will be sampled at 30Hz because that is the sampling rate of the camera. In reality, the AFFDEX engine live processes data at every other sample, or frame, to speed up processing. So, for a 30Hz camera, AFFDEX will live process at 15Hz.

What this means is that when you are done collecting data for a subject, you may get red bars around the quality assurance metric for AFFDEX, as pictured below:



This means that iMotions was expecting AFFDEX to pick up at 30Hz instead of 15Hz. If this value is between 40-50%, this is not cause for concern. You can post process to analyze the skipped video frames. How to Post-Process Videos using AFFDEX.

It is always good practice to review the replay to check for any possible sources of error (i.e. subject turns away from the camera, missing data), especially if the data quality is less than 40%.

If you have further questions about this, please contact support@imotions.com

Powered by Document360

Confidential © 2023 Redistribution is not permitted without written permission from iMotions

Confidential © 2023 Redistribution is not permitted without written permission from iMotions