

Does the ECG channel help? ECG+PPG vs PPG-only

Model	ECG + PPG		PPG only	
	SBP	DBP	SBP	DBP
LeNet1d	12.7	8.5	13.6	8.6
Inception1d	12.9	8.6	13.3	8.0
XResNet50	12.8	8.7	13.2	8.4
XResNet101	12.8	8.6	13.2	8.2
Transformer	13.3	8.6	13.4	8.2
LightGBM	12.7	8.1	12.7	8.1

MAE (mmHg), in-distribution. Removing ECG costs <0.5 mmHg -- the arrival-time channel adds little, consistent with the roll-audit showing models barely use PAT.