





b Datasets

Dataset	Role	Seg	Subj	SBP/DBP	Age [5-95]	Chan
VitalDB	train	465k	1,293	115/63	61 [31-79]	ECG+PPG
VitalDB test	ID	58k	144	115/63	60 [30-79]	ECG+PPG
MIMIC-BP	OOD	137k	1,524	113/58	--	ECG+PPG
BCG	OOD	3k	40	121/67	28 [19-65]	PPG
Sensors	OOD	11k	1,195	134/65	57 [33-81]	PPG
UCI2	OOD	411k	10,793	132/67	--	PPG
PPG-BP	OOD	619	218	128/72	58 [25-80]	PPG



d DBP MAE (mmHg), size + roll-audit mechanism

Model	#par	ID	OOD	slope	rank
LeNet1d	659k	8.5	13.5	-7	4
Inception1d	472k	8.6	11.8	-20	2
XResNet50	887k	8.7	12.6	-28	1
XResNet101	1.8M	8.6	12.5	+9	3
Transformer	107k	8.6	13.6	-1	5
LightGBM	~8k	8.2	15.1	--	--

