

Model	Accuracy (DBP MAE)		Roll-audit (relative)		
	ID (mmHg)	MIMIC (OOD) (mmHg)	slope (mmHg/s)	rank (sensitivity)	response range
XResNet1d50	8.69	12.56	−28.2	1 (most)	large
Inception1d	8.57	11.80	−19.5	2	mod.
LeNet1d	8.45	13.54	−7.3	3	small
Transformer	8.57	13.58	−1.4	4 (least)	flat
XResNet1d101	8.61	12.52	+8.6	opposite	large

The roll-audit measures each model’s relative sensitivity to an arrival-time perturbation (identical across models). The cross-model ranking and OOD correlation are meaningful; the absolute slope magnitude is not physiologically calibrated and is not interpreted here.