Type of testing and measuring instrument	producer	machine / serial / model number	measuring range	validation / calibration measurand	(international) standard	class	validation / calibration information	operator w/ DAkkS accreditation
charpy test machine	Zwick/Roell	type 750 J	energy: up to 750 J	energy			calibration ID: H70388, D-K-18351-01-00 calibration date: 2021-01	service provider
dial gauge	Mitutoyo	2046F-09 | 46K303	0 - 30 mm (0.01 mm)	test piece diameter	DAkkS-DKD-R 4-3 sheet 10.1	10 µm	calibration ID: 47073, D-K-17059-02-00 calibration date: 2021-10	service provider
caliper	Mitutoyo	Absolute Coolant Proof IP 67	0 - 150 mm	test piece diameter	DAkkS-DKD-R 4-3 sheet 9.1:2010		calibration ID: 47054, D-K-17059-02-00 calibration date: 2021-10	service provider