Type of testing and measuring instrument	**producer**	**machine / serial / model number**	**measuring range**	**validation / calibration measurand**	**(international) standard**	**class**	**validation / calibration information**	**operator w/ DAkkS accreditation**
electronic weighing scale	Sartorius AG	BP210D / 60403839	max. 80 g (0.00001 g) max. 200 g (0.0001 g)	weight	Euramet/cg-18/V.4.0	OIML-class E2	calibration ID: F-0294, D-K-18186-01-00 calibration date: 2021-10-20	service provider
dial gauge	Vogel Germany GmbH & Co. KG	140219042	0 - 25 mm (0.001 mm)	test piece diameter	DAkkS-DKD-R 4-3 sheet 10.1:2018		calibration ID: 019703, D-K-15048-01-00 calibration date: 2021-03-19	service provider
caliper	Mitutoyo	type 500-312 / 7015071	0 - 200 mm (0.01 mm)	test piece diameter	DKD-R 4-3 Blatt 9.1:2018		calibration ID: 019591, D-K-15048-01-00 calibration date: 2021-03-19	service provider
network analyzer	Hewlett Packard	HP 8751 A / 3315J01624	100 Hz - 100 kHz	frequency, stability, starting behavior of the time basis	EA-4/02 M: 2013		calibration ID: 18151, D-K-17543-01-00 calibration date: 2021-06-17	service provider