PRESSURE MEASUREMENT

Product information pack

Anhui Jingfeng Instrument Co., Ltd. | English review edition

Not a released datasheet. This pack compiles current website descriptions and sample information. Numerical values, materials, compatibility and supply capability have not been independently manufacturer-verified.

Four product records, one selection workflow

JF-1541	Pressure Gauge | Bourdon-tube description
JF-1545	Stainless Steel Pressure Gauge | flange sample
JF-1543	Capsule Pressure Gauge | low-pressure gas description
JF-1542	Ammonia Pressure Gauge | dual-scale sample

How to use this pack

1. Identify the instrument family using the sample summaries.
2. Complete the inquiry worksheet with your actual operating conditions.
3. Obtain a manufacturer-approved specification and dimensional drawing.
4. Release the final purchase specification only after engineering review.

Source: repository `src/lib/products.ts` and existing product images. The site model identifiers are reproduced for inquiry matching. No order code, product certification, pressure rating or performance guarantee is established by this draft.

SAMPLE PRODUCT RECORDS

Identify the instrument

All values below are existing website/sample information, not a verified supply specification.

JF-1541 | Pressure Gauge



Reported sample: 0-1.6 MPa; accuracy class 1.6

Bourdon-tube mechanical pointer indication is described on the website.

Selection limits: Case described as black and circular. Thread, diameter, wetted materials and temperature limits are not confirmed.

Website reference: /en/products/pressure-gauge

JF-1545 | Stainless Steel Pressure Gauge



Reported sample: 0-2.5 MPa; accuracy class 1.6

Mechanical pressure gauge with a flange assembly shown.

Selection limits: The website reports a 316L marking on the lower connector. This does not establish the alloy of all wetted parts, flange standard or pressure rating.

Website reference: /en/products/stainless-pressure-gauge

SAMPLE PRODUCT RECORDS

Identify the instrument

All values below are existing website/sample information, not a verified supply specification.

JF-1543 | Capsule Pressure Gauge




Reported sample: 0-10,000 Pa; accuracy class 2.5

Capsule-element local indication for low gas pressure is described.

Selection limits: Side threaded connection is reported. Thread standard, element material, gas compatibility and overload limit require confirmation. Image is an existing edited website asset.

Website reference: /en/products/capsule-pressure-gauge

JF-1542 | Ammonia Pressure Gauge



Reported sample: 0-4 MPa / 0-580 psi; accuracy class 1.6

Mechanical dual-scale indication with a bottom connection shown.

Selection limits: The product name is not proof of compatibility. Confirm exact refrigerant, all wetted materials, seals, pressure and temperature ratings before ammonia service.

Website reference: /en/products/ammonia-pressure-gauge

INQUIRY SPECIFICATION WORKSHEET

Define the application

Complete one worksheet per duty. Attach the process specification and installation drawing where available.

Project / destination	_____
Model / quantity	_____
Medium / composition	_____
Normal / min / max pressure	_____
Pressure unit / gauge type	_____
Process / ambient temperature	_____
Vibration / pulsation / shock	_____
Connection / mounting	Thread standard, size, male/female OR flange standard and rating: _____
Dial / orientation / space	_____
Material / seal requirements	_____
Documents / inspection	Required drawing, calibration, inspection or material records: _____

Also provide required delivery date, packaging and commercial terms separately. Submission of requirements does not confirm availability or acceptance. Never substitute a visually similar thread or infer material compatibility from case appearance.

RELEASE CHECKLIST

From draft to datasheet

Engineering and business owners must complete the evidence review before this pack is published or sent as an approved catalog.

Product identity	Confirm each model against a current manufacturer record.
Measurement specification	Approve range options, accuracy basis, operating limits and overload conditions for each model.
Materials and compatibility	Record case, sensing element, connector and seal materials separately; confirm application suitability.
Installation	Supply dimensioned drawings, connection standards, sizes, orientation and mounting clearances.
Options and documentation	Confirm actual accessories and available inspection/calibration records. Do not imply unavailable certifications.
Photos and commercial terms	Approve images and product labels; verify MOQ, lead time, packing and export information.
Approval and release	Name the engineering reviewer, evidence references, review date and revision. Publish only the approved revision.

Engineering reviewer: Not assigned
Evidence files: Not supplied
Approval date: Pending
Release state: Internal draft; not for procurement specification or installation.