

Let there be N measurements M_{meas} , with each measured point being approximated well by a fitted value M_{fit} dependant on L fitting or/and input variables φ :

$$\begin{bmatrix} M_{1,fit}(\varphi_1, \dots, \varphi_L) \\ \vdots \\ M_{N,fit}(\varphi_1, \dots, \varphi_L) \end{bmatrix} \cong \begin{bmatrix} M_{1,meas} \\ \vdots \\ M_{N,meas} \end{bmatrix}$$

In the case of a small error in a measured input variable $\varphi_{j,meas}$, the change in a fitted value of M_i can be approximated by a Taylor expansion:

$$M_{i,fit}(\varphi_{j,meas} + \Delta\varphi_{j,meas}) = M_{i,fit}(\varphi_{j,meas}) + \sum_{k=1}^L \frac{\partial M_{i,fit}}{\partial \varphi_k} \Delta\varphi_k + \dots$$

As we always fit the model to the original fitted measurement values, and we assume a good fit:

$$M_{i,fit}(\varphi_{j,meas} + \Delta\varphi_{j,meas}) \cong M_{i,meas} \equiv M_i$$

and

$$M_{i,fit}(\varphi_{j,meas}) \cong M_{i,meas} \equiv M_i$$

such that

$$M_i \cong M_i + \sum_{k=1}^L \frac{\partial M_i}{\partial \varphi_k} \Delta\varphi_k$$

Thus, any error caused by $\Delta\varphi_{j,meas}$ is compensated by variations in the other fitted parameters $\Delta\varphi_k$:

$$\sum_{\substack{k=1 \\ k \neq j}}^L \frac{\partial M_i}{\partial \varphi_k} \Delta\varphi_k + \frac{\partial M_i}{\partial \varphi_{j,meas}} \Delta\varphi_{j,meas} = 0$$

$$\sum_{\substack{k=1 \\ k \neq j}}^L \frac{\partial M_i}{\partial \varphi_k} \Delta\varphi_k = - \frac{\partial M_i}{\partial \varphi_{j,meas}} \Delta\varphi_{j,meas}$$

For qMT, the measurement we are interested in as a possible source of error is B_1 , and the fitted parameters are F , k_f , T_{2f} and T_{2r} ($T_{1,f}$ is omitted, it's simply a function of $T_{1,meas}$, F , and k_f)

$$\frac{\partial M_i}{\partial F} \Delta F + \frac{\partial M_i}{\partial k_f} \Delta k_f + \frac{\partial M_i}{\partial T_{2,f}} \Delta T_{2,f} + \frac{\partial M_i}{\partial T_{2,r}} \Delta T_{2,r} = - \frac{\partial M_i}{\partial B_1} \Delta B_1$$

The sensitivity function is defined as:

$$S_{\varphi_k,i} \equiv \frac{\partial M_i}{\partial \varphi_k}$$

Such that the previous equation can be writing compactly as:

$$\begin{bmatrix} S_{F,i} & S_{k_f,i} & S_{T_{2,f},i} & S_{T_{2,r},i} \end{bmatrix} \begin{bmatrix} \Delta F \\ \Delta k_f \\ \Delta T_{2,f} \\ \Delta T_{2,r} \end{bmatrix} = -S_{B_1,i} \Delta B_1$$

From this equation, we can provide the following general observations:

1. The larger the sensitivity functions $S_{\varphi_k,i}$, the less variation $\Delta\varphi_k$ (e.g. $\Delta F, \Delta k_f, \dots$) is required for φ_k (e.g. F, k_f, \dots) to compensate the error caused by $\Delta\varphi_{j,meas}$ (e.g. ΔB_1).
2. The smaller the sensitivity function for the measurement $S_{\varphi_k,i}$ (e.g. $S_{B_1,i} = \frac{\partial M_i}{\partial B_1}$), the less overall compensation is required by all the φ_k (e.g. F, k_f, \dots).
3. $S_{B_1,i}$ will depend on the T_1 mapping method (e.g. $S_{B_1,i}^{VFA}, S_{B_1,i}^{IR}$), due to the B_1 dependence of VFA and B_1 independence of IR T_1 maps.

For a set of N measurements (for our qMT protocols, $N=10$), this equation expands to:

$$\begin{bmatrix} S_{F,1} & S_{k_f,1} & S_{T_{2,f},1} & S_{T_{2,r},1} \\ \vdots & \vdots & \vdots & \vdots \\ S_{F,N} & S_{k_f,N} & S_{T_{2,f},N} & S_{T_{2,r},N} \end{bmatrix} \begin{bmatrix} \Delta F \\ \Delta k_f \\ \Delta T_{2,f} \\ \Delta T_{2,r} \end{bmatrix} = \begin{bmatrix} -S_{B_1,1} \Delta B_1 \\ \vdots \\ -S_{B_1,N} \Delta B_1 \end{bmatrix}$$

Which can be written in matrix form:

$$S\psi = \Psi$$

where S is the matrix of sensitivity functions, ψ is the vector of changes in fitted parameters, and Ψ is the vector of $-S_{B_1,i} \Delta B_1$.

As the equation is an overdetermined set of linear equations of the form:

$$Ax = b$$

The optimal changes in parameters ψ can be solved through the following minimization:

$$\min \|S\psi - \Psi\|_2^2$$

for different protocols, B_1 mapping methods, etc. S and Ψ can be calculated analytically (possible, but cumbersome to do so), or estimated through simulations.

From this analysis, and from my qMT B1-insensitivity measurements, we can make the following general predictions:

- $|S_{F,i}| \gg |S_{k_f,i}|$ for most measurements (if not all), since $\Delta F \cong 0$ and Δk_f
- $|S_{k_f,i}| \ll |S_{F,i}|, |S_{T_{2,f},i}|, |S_{T_{2,r},i}|$ for most measurements (if not all) (most of the error is dumped into kf)
- $|S_{B_1,i}^{VFA}| < |S_{B_1,i}^{IR}|$ since less ΔF is required for VFA and IR. (This one is less strong, since Δk_f is smaller for IR and VFA, but for ΔF it's the opposite. Possibly a more complicated relationship to the solution of minimization instead of this simple relation)