


# A B<sub>1</sub>-Insensitive qMT Protocol



Mathieu Boudreau<sup>1</sup>, Nikola Stikov<sup>1</sup>, G. Bruce Pike<sup>2</sup>

<sup>1</sup>McConnell Brain Imaging Center, Montreal Neurological Institute, McGill University, Montreal, Quebec, <sup>2</sup>Hotchkiss Brain Institute, Faculty of Medicine, University of Calgary, Calgary, Alberta

## Introduction

- Quantitative magnetization transfer (qMT) imaging requires  $B_0$  and  $B_1$  measurements to correct for instrumental biases, and a  $T_1$  measurement to constrain parameters in the fitting model.
- If using Variable Flip Angle (VFA) T<sub>1</sub> mapping<sup>1</sup>, B<sub>1</sub> is used twice before fitting the qMT parameters: to correct the flip angles for T<sub>1</sub> mapping, and to scale the nominal MT saturation powers.
- Inaccuracies in B<sub>I</sub> would propagate to the fitting of the qMT parameters through two pathways – through errors induced in T<sub>I</sub>, and errors in MT saturation powers.
- This work demonstrates that for the Sled and Pike qMT model<sup>2</sup>, certain qMT parameters are insensitive to a large range of B<sub>1</sub> inaccuracies when using VFA for T<sub>1</sub> mapping.

## Methods

- Siemens 3T Tim Trio MRI system, 32-channel head coil
- 3 healthy adult volunteers
- Single slice, AC-PC orientation, slightly above the corpus callosum (2x2x5 mm<sup>3</sup>)

### **Pulse Sequences**

- $B_0 \rightarrow Two$ -point GRE phase-difference method
- $B_1 \rightarrow Double angle method (DAM) (<math>\alpha = 60^{\circ}/120^{\circ})$
- $T_1 \rightarrow Variable Flip Angle^3$  (TR = 15 ms,  $\alpha = 3^{\circ}/20^{\circ}$ )  $\rightarrow Inversion recovery^{4,5}$  (TI = 30, 530, 1030, 1530 ms)
- qMT → Spoiled GRE optimal 10-point protocol<sup>6</sup>,
  Gaussian-Hanning MT pulses, Sled and Pike qMT model<sup>5</sup>

### Simulated B<sub>1</sub> Errors

•  $B_1 \rightarrow Flat B_1 maps = 0.5, 0.75, 0.9, 1, 1.1, 1.25, 1.5, 2 n.u.$ 





Figure 1. Quantitative MRI protocol processing hierarchy.

# Results – Measured B, vs. Nominal FA




**Figure 2.** Comparison of VFA  $T_1$  and qMT F maps using measured (DA) and nominal ( $B_1$  flat = I)  $B_1$  maps.

# Results - VFA vs. IR

Figure 3 shows the pooled whole brain Pearson correlation coefficients (a) and linear regression slopes (b) for qMT F values between the measured DA  $B_1$  maps and simulated flat  $B_1$  maps, for VFA (blue) and IR (red)  $T_1$  maps.


High correlation (a) and linear regression slope values near I (b) for qMT F values are observed using VFA  $T_1$  maps for a large range of flat  $B_1$  maps (0.75 - 2 n.u.). qMT F maps fitted using IR  $T_1$  are sensitive  $B_1$  errors, as expected.



**Figure 3.** Linear regression analysis of the voxelwise qMT F parameter, comparing measured DA  $B_1$  and a range of flat  $B_1$  (VFA  $T_1$ , blue; IR  $T_1$ , red).

# Results – qMT Parameters

**Table 1.** Voxelwise linear regression analysis of all fitted qMT parameters, using each  $T_1$  method (VFA - left, IR-right), and a comparison of the measured (DA) and nominal FA ( $B_1$  flat = 1) maps.



qMT	Pearson ρ	Slope	Pearson ρ	Slope
F	0.99	0.98	0.81	0.81
$\mathbf{k_f}$	0.32	0.31	0.52	0.57
$R_{1f}$	0.81	0.98	0.78	0.71
$T_{2f}$	0.99	0.95	0.93	1.02
$T_{2r}$	0.92	0.90	0.87	0.91

## Discussion

- VFA-based T<sub>1</sub> maps renders qMT F and T<sub>2,f</sub> insensitive to B<sub>1</sub> errors.
- Processing qMT F maps using VFA T<sub>1</sub> and a flat B<sub>1</sub> map (nominal flip angle assumption, large B<sub>1</sub> inaccuracies) results in nearly identical qMT F maps using DA B<sub>1</sub> maps, (Fig. 2), except where CSF partial volume effects are suspected.
- Severe overestimation of B<sub>I</sub> is better tolerated than severe underestimation for the qMT parameter F (Fig. 3).
- The exact origin of the erroneous  $B_1$  and VFA  $T_1$  nearly cancelling out in qMT F maps remains to be clarified.
- A possible explanation might be that errors in B1 propagate to F via counterbalancing effects on T1 estimation and MT saturation power

## Summary

- This work demonstrated that qMT F maps fitted using VFA T<sub>1</sub> can be insensitive to B<sub>1</sub> inaccuracies.
- A strong correlation (0.99) between qMT F values fitted using measured and nominal B<sub>I</sub> maps was observed when using VFA T<sub>I</sub>
- More work in simulating the effects of B<sub>1</sub> and VFA T<sub>1</sub> inaccuracies on qMT parameter estimation is needed to have a clearer understanding of the limitations of this observation.

**References:** [1] Deoni S. et al, MRM 49:515-526 (2003) [2] Sled J. and Pike G. B., MRM 46:923-931 (2001) [3] Yarnykh V., MRM 63:1610-26 (2010) [4] Barral J. et al, MRM 64:1057-1067 (2010) [5] http://www-mrsrl.stanford.edu/~jbarral/t1map.html (Accessed: October 2012) [6] Levesque I. et al, MRM 66:635-643 (2011) [7] Schmierer K. et al, JMRI 26:41-51 (2007) [8] Yarnykh V., MRM 68:166-178 (2012)

**Acknowledgments:** This work was funded by the Natural Sciences and Engineering Research Council's Doctoral Alexander Graham Bell Canada Graduate Scholarship (M.B.), and grant funding was provided by the Canadian Institutes of Health Research.

Contact: mathieu.boudreau2@mail.mcgill.ca