

From: Simon Wilkinson simon.wilkinson@tanita.eu
Subject: Tanita TBF 305
Date: 8 April 2020 at 09:30
To: matthew.lee@bristol.ac.uk



Good morning Matt,

Apologies nobody has come back to you on this, like most places we are down to a skeleton staff at HQ in Amsterdam.

We have received communication from Tokyo HQ that unfortunately it is not possible to convert the impedance values only into a body fat percentage from a manual input, this would have required software to be utilised at the point of measurement.

If there is anything else I might be able to help with, just let me know.

--

Healthy regards

Simon Wilkinson
Technical Rep Fitness & Sport

TANITA

FIT STARTS WITHIN

Tanita Europe BV (UK Office)
111 Piccadilly
Manchester
M1 2HY

Tel: +44(0)161 6380926
Mob: +44(0)7795278733

**STAY HOME.
STAY HEALTHY.**



www.tanita.eu