

# A CRYOGENIC TESTING ENVIRONMENT FOR SPT-3G, THE NEXT GENERATION SOUTH POLE TELESCOPE RECEIVER

MATTHEW YOUNG

Astronomy & Astrophysics Department, University of Toronto, ON

SUPERVISED BY: KEITH VANDERLINDE AND TYLER NATOLI

50 St. George Street  
 Toronto, Ontario, Canada  
 M5S 3H4

*Draft version April 17, 2016*

## ABSTRACT

The next generation optical system for the South Pole Telescope, SPT-3G, is set to be installed in early 2016. This report outlines the fundamental science goals and operation behind SPT-3G, and covers the continued development of a cryogenic testing environment at the University of Toronto for characterizing the next generation of detectors.

### 1. INTRODUCTION

The South Pole Telescope (SPT) is a 10 metre microwave telescope, located at the geographic South Pole in Antarctica. Designed to observe temperature and polarization anisotropies in the Cosmic Microwave Background (CMB) (Ruhl et al. 2004), the telescope has been optimized to operate in the millimetre (mm) band spanning from 75 GHz to 240 GHz. The next generation optical system to be installed at the beginning of 2016, SPT-3G, is to contain over 16,000 polarization sensitive detectors in the focal plane array. This order of magnitude increase in the number of detectors over the current receiver will allow for high signal-to-noise mapping of both the *E*-mode and *B*-mode CMB polarization anisotropies (Benson et al. 2014). With the increased sensitivity achieved by SPT-3G, significant advances can be made in the fields of large scale structure formation, particle physics, and inflation.

The SPT-3G detector wafers are currently being fabricated by Argonne National Laboratory, using a range of designs in order to tweak device parameters. Each wafer is tested within the SPT collaboration, providing characteristic feedback for the nanofabrication process. The detectors operate in the superconducting transition for niobium, requiring a sub-kelvin cryogenic environment for detector testing. This report outlines the development of such a system at the University of Toronto (UofT), to assist in the testing and characterization of SPT-3G wafers. In conjunction, this system also features the SPT-3G Digital Frequency Multiplexing (DfMUX) readout system. This enables testing of the entire SPT-3G detector readout chain at UofT. In Section 2, the science goals of SPT-3G are outlined, with the instrumentation required to achieve these goals covered in Section 3. Section 4 details the development of the cryogenic testing at UofT, with current testing results in Section 5. Finally, Section 6 takes a brief look at processing SPTpol detector timestreams into usable scientific data.

### 2. SPT-3G SCIENCE GOALS

The telescope is located at the geographical South Pole, providing one of best sites on Earth for observing

mm-wavelength anisotropies. This is in part due to the South Pole's elevation of 2.8 kilometres above sea level, resulting in desirably low levels of atmospheric fluctuation power. The near zero air humidity also weakens the strong absorption and emission features of atmospheric water vapour at mm-wavelengths. The current optical design produces a beamwidth of  $\sim 1$  arcmin at 150 GHz, allowing for high resolution mapping of small-scale CMB features. Coupled with the high raw sensitivity achievable through the new multi-chroic polarization-sensitive pixel design, SPT-3G will be able to extract a wealth of previously unobtainable cosmological information from the CMB.

Lensing measurements of the CMB are capable of probing matter fluctuations out to the last scattering surface, providing a detailed lensing power spectrum. SPT-3G's precise measurement of the growth of structure on small angular scales places tight constraints on the number of relativistic species at recombination. Combined with measurements from the *Planck* experiment, it is expected that the sum of neutrino masses will be constrained to within  $\sigma(\Sigma m_\nu) \sim 0.06$  eV (Benson et al. 2014).

### UP TO HERE

### 3. SPT-3G INSTRUMENTATION

#### 3.1. Focal Plane Array

Each SPT-3G wafer fabricated by Argonne National Laboratory contains 217 pixels, with each pixel featuring a sinuous log-periodic antenna, microstrip inductor-capacitor (LC) filters, and six transition edge superconducting (TES) bolometers (Benson et al. 2014). A photograph of the pixel design is shown in Figure. The bolometers are responsible for detecting photons at band centres of either 95 GHz, 150 GHz, or 220 GHz for a given polarization, following the isolation of each signal by the LC filters. As the TES bolometers are maintained in a superconducting phase transition through electro-thermal feedback (Benson et al. 2014), a small increase in temperature due to the incident photons will result in a large change in resistivity.

#### 3.2. Readout System

Reading out the thermal loading of each bolometer individually introduces an unnecessary number of wires to the sub-Kelvin stage, and as such, SPT-3G utilizes a frequency multiplexing readout system, enabling 64 bolometers to be addressed using a single wire (Dobbs et al. 2012). By placing each resistive bolometer in series with a unique LC filter, an LCR resonator is formed (Henning et al. 2012) that can be interrogated using its characteristic resonant frequency, allowing the warm electronics to perform readout in Fourier space. A digital frequency multiplexing (DfMUX) (Dobbs et al. 2012) system has been developed by McGill University, designed to deliver small amplitude excitations to each resonator with active digital feedback for nulling the subsequent signal. The error in signal cancellation is then sensitively measured using superconducting quantum interference devices (SQUIDs) (Dobbs et al. 2012), inferring changes in the bolometer impedance and hence thermal loading.

#### 4. TESTING ENVIRONMENT

The Long Wavelength Lab currently houses an Olympus 104 cryostat, designed and fabricated by *High Precision Devices* (HPD) in Colorado. Featuring a helium pulse tube refrigerator in conjunction with an adiabatic demagnetization refrigeration (ADR) system, the cryostat is capable of achieving temperatures as low as 35 mK. These sub-Kelvin temperatures are realized through the ADR, consisting of two paramagnetic salt pills within a superconducting electromagnet, backed by the 3 K base temperature of the pulse tube. By applying a large current to the electromagnet, magnetic domains within the salt pills will begin to align, freezing out a thermal degree of freedom within the salt pills. Once fully magnetized, the current can be removed, allowing

magnetic domains to realign themselves and reintroducing the thermal degree of freedom. In accordance with the conservation of energy, the realignment process draws heat from the system, lowering the pills to sub-Kelvin temperatures.

The cryogenic testing environment also features an interactive control system and web interface, facilitating the measurement of system parameters such as temperature, pressure, and electromagnet current, as well as conducting the magnetization process. An image of the current web interface is shown in Figure. Following demagnetization, a proportional-integral-derivative (PID) controller maintains the cold-stage temperature at a predetermined setpoint, capable of regulating temperatures of 100 mK for 100 hours.

#### 5. TESTING RESULTS

SQUIDS 4 dayz.

#### 6. DATA PROCESSING

Looking towards the future!

#### REFERENCES

- Addison, G. E., Huang, Y., Watts, D. J., et al. 2015, arXiv:1511.00055 [astro-ph]
- Samtleben, D., Staggs, S., Winstein, B. 2007, in Annual Review of Nuclear and Particle Science, 245-283
- Bender, A. N., Cliche, J. F., Haan, T., et al. 2014, arXiv:1407.3161 [astro-ph]
- Henning, J. W., Ade, P., Aird, K. A., et al. 2012, arXiv:1210.4969 [astro-ph]
- Dobbs, M. A., Lueker, M., Aird, K. A., et al. 2012, Review of Scientific Instruments, 83, 7
- Benson, B. A., Ade, P. A. R., Ahmed, Z., et al. 2014, arXiv:1407.2973 [astro-ph]
- Ruhl, J. E., Ade, P. A. R., Carlstrom, J. E., et al. 2004, arXiv:astro-ph/0411122