

CERTIFICATE OF ANALYSIS

Precision Plastics Corp

ISO 13485:2016 Certified | FDA Registered Facility

Product	1-Way Stopcock, Female Luer Lock, Male Luer Lock
Part Number	SP-11195-A (Qosina equivalent: #11195)
Lot Number	LOT-2026-0501
Mfg Date	March 15, 2026
Expiry Date	March 15, 2031 (60 months)
Quantity	10,000 units

TEST RESULTS

Test	Specification	Result	Status
Visual Inspection	No defects	No defects	PASS
Thru-Hole Diameter	2.69mm +/- 0.05	2.70mm	PASS
Burst Pressure	> 29 psi	42 psi	PASS
Handle Torque	> 3 in-oz	4.2 in-oz	PASS
Biocompatibility	ISO 10993-1	Compliant	PASS
Particulate Matter	< 50 particles/mL	12 particles/mL	PASS
Endotoxin	< 20 EU/device	< 0.5 EU/device	PASS
Sterility (pre-irrad)	SAL 10 ⁻⁶	SAL 10 ⁻⁶	PASS

MATERIALS:

Body: Polycarbonate (PC) | Handle: HDPE | Seal: Silicone
Connections: Female Luer Lock inlet, Male Luer Lock outlet
Compliance: ISO 80369-7

This certifies that the above lot has been manufactured and tested in accordance with applicable specifications and is released for distribution.

QA Manager: Dr. Li Wei | Date: March 20, 2026