

Relatorio Tecnico SGMed Inspector

Escopo: 21JN101B

Resumo Executivo

Mudancas encontradas: 5 | Eventos avaliados: 2788

O Que Mudou

Parametro	Esquerda	Direita	Categoria	Severidade
Date / time	-	15/04/26 06:10:25	network	medium
Date / time	14/04/26 06:15:00	-	operational	medium
Standard dens override	854.4	851.4	metrological	high
Override	854.4	851.4	metrological	critical
Override	854.4	851.4	metrological	critical

Cromatografia / GC

Componentes com mudança: nenhum componente com mudança

Componentes lidos no arquivo carregado: nenhum componente lido

Perfil esperado: Metano, Etano, Propano, Hexano+, Nitrogênio, CO2

Componentes esperados ainda ausentes: Metano, Etano, Propano, Hexano+, Nitrogênio, CO2

Leitura dos Eventos

Run 1 Flow rate alarm from High High to Normal [170.17] (902x) - Run 1

Run 1 Flow rate alarm from Normal to High High [170.17] (879x) - Run 1

Run 2 Flow rate alarm from Normal to Low Low [80] (480x) - Run 2

Run 2 Flow rate alarm from Low Low to Normal [80] (458x) - Run 2

User administrator (Web: 10.0.1.50) has logged in (17x) - Sem run especifico

Run 1 Flow rate alarm from High High to Normal [170.17] (742x) - Repeticao em janela curta no run 1

Run 1 Flow rate alarm from Normal to High High [170.17] (728x) - Repeticao em janela curta no run 1

Run 2 Flow rate alarm from Normal to Low Low [80] (429x) - Repeticao em janela curta no run 2

Run 2 Flow rate alarm from Low Low to Normal [80] (409x) - Repeticao em janela curta no run 2

User administrator (Web: 10.0.1.50) has logged in (8x) - Repeticao em janela curta