Made-up QA/QI Plan

Made-up Plan Print Lab | CLIA: 22D7790015 | Effective 2027 | Printed 2026-11-12

Program scope

Both made-up sites, all sections.

Measurement methods

Monthly counts from saved reports.

Analysis methods

Each month against its goal.

Remediation methods

An action plan when a goal is missed.

Monitor and sustain methods

Three months at goal before closing.

Quality indicators (3)

San Carlos: Chemistry (2)

Indicator	Unit	Goal	How it's measured
Hemolysis rate	%	≤ 2	Hemolyzed rejections / chemistry specimens received.
STAT potassium TAT	%	≥ 90; within 45 min	Meditech TAT report.

Bylas: Hematology (1)

Indicator	Unit	Goal	How it's measured
Clotted CBCs	%	≤ 1; yellow ≤ 2	Clotted CBC rejections / CBCs received.

Improvement priorities

Critical value calls
Goal: Within 30 minutes
Department: San Carlos: Chemistry

Annual leadership review

Date	Reviewed by	Title	Changes	Next due
2026-11-01	Made-up Director	Medical Director	Adopted for 2027.	2027-11-01