Instructions for use: NeuroStat Diadem

NeuroStat Diadem

Read all of this document carefully before you start. Keep it: you may need to read it again. If you have further questions, ask your doctor, pharmacist or occupational health service.

1. What NeuroStat Diadem is

NeuroStat Diadem is an optically pumped magnetometer array, class IIa. It is intended for: recording of brain activity during movement and sleep, in clinical neurophysiology and research.

How it works

Diadem is a helmet that holds 64 optically pumped magnetometers. They pick up the very weak magnetic fields that active neurons produce. Unlike the sensors of a conventional MEG system, these work at room temperature and sit a few millimetres from the scalp, so the helmet weighs under 1.5 kilograms and can be worn standing, walking or lying down. Coils built into the room cancel the earth's magnetic field. The Diadem workstation shows when and where activity happens.

2. Before you use it

- Diadem must be used in the magnetically controlled room installed by NeuroStat Diagnostics. Do not move the coils or the shielding.
- Take off watches, jewellery and other metal, and switch off phones, before a recording. They disturb the measurement.
- The sensor housings are sealed. Do not open them.
- Not recommended for people with pacemakers or neurostimulators unless the treating physician agrees.

3. Reporting problems

If you notice any side effect or problem, talk to your doctor, pharmacist or occupational health service. You can also report it to your national authority and to NeuroStat at safety@neurostat.com. Reporting helps provide more information on the safety of this product.

4. About this document

Manufacturer: NeuroStat Diagnostics AG, Basel, a company of the NeuroStat group. Marketing responsibility: NeuroStat Holding AG, Thurgauerstrasse 66, 8050 Zürich, Switzerland. Launched 2028. Last revised September 2028.