Instructions for use: NeuroStat Tenor

NeuroStat Tenor

Read all of this document carefully before you start. Keep it: you may need to read it again. If you have further questions, ask your doctor, pharmacist or occupational health service.

1. What NeuroStat Tenor is

NeuroStat Tenor is a transcranial ultrasound neuromodulation system, research use. It is intended for: research into how stimulating a small region of the brain changes sleep, attention and memory.

How it works

Tenor sends low-intensity ultrasound through the skull to a spot a few millimetres wide. At these levels the ultrasound does not heat or damage tissue. It changes how easily the neurons in the target fire, for minutes to hours. A transducer array in a head frame aims the beam. Tenor works inside a 3 tesla MRI scanner, so researchers can see the target and the response in the same session.

2. Before you use it

- Tenor is for research use only. It is not approved to treat any condition.
- Studies need ethics approval. Participants are screened for epilepsy, implants and pregnancy.
- Mild headache, tingling and neck discomfort have been reported. They pass within hours.
- The beam is checked against the safety limits before every session. Tenor will not start if a reading is out of range.

3. Reporting problems

If you notice any side effect or problem, talk to your doctor, pharmacist or occupational health service. You can also report it to your national authority and to NeuroStat at safety@neurostat.com. Reporting helps provide more information on the safety of this product.

4. About this document

Manufacturer: NeuroStat Diagnostics AG, Basel, a company of the NeuroStat group. Marketing responsibility: NeuroStat Holding AG, Thurgauerstrasse 66, 8050 Zürich, Switzerland. Launched 2028. Last revised September 2028.