Instructions for use: NeuroStat Halo

NeuroStat Halo

Read all of this document carefully before you start. Keep it: you may need to read it again. If you have further questions, ask your doctor, pharmacist or occupational health service.

1. What NeuroStat Halo is

NeuroStat Halo is wearable EEG device, class IIa. It is intended for: sleep staging and fatigue monitoring in shift-working populations.

How it works

Halo records frontal EEG through dry electrodes, with no gel and no preparation. Software on the device classifies sleep stages and estimates fatigue from changes in spectral power and sleep architecture. Occupational health teams review the summaries on a clinician dashboard.

2. Before you use it

- Intended for use under the supervision of a healthcare professional or occupational health service.
- Halo is not meant for diagnosing epilepsy or other seizure disorders.
- Skin irritation can occur where the electrodes touch. Stop using the device if it does not go away.
- Do not use with implanted devices in the head unless the treating clinician agrees.

3. Reporting problems

If you notice any side effect or problem, talk to your doctor, pharmacist or occupational health service. You can also report it to your national authority and to NeuroStat at safety@neurostat.ch. Reporting helps provide more information on the safety of this product.

4. About this document

Marketing authorisation holder: NeuroStat Holding AG, Thurgauerstrasse 66, 8050 Zürich, Switzerland. First approved 2019. Last revised September 2028.