MASTER COPY

Biomedical Limited, No 1 Ohimege Road, Industrial Estate, Ilorin Kwara State, PMB 1449.

DEMINERALISED WATER ANALYSIS RECORD BOOK

SAMPLING POINT: DM Storage tank

REFERENCE: BP2024

QC AR NO: QC/DM/25/SYP/

FORM NO: DMR003

Date:

Time:

Parameters	Limits	Results
Description	Colourless, odourless, tasteless	
pH	5.5 – 7.5	
Conductivity	≤ 5.1µs/cm @ 25^{0}C	
Acidity/Alkalinity	NMT 0.5mL of 0.01M NaOH/HCl is required to change the solution to yellow/red.	
Total Hardness	NMT 5ppm	
Chlorides	≤ 0.2ppm	
Sulphates	The turbidity produced in test solution for at least 1 hour should not be greater than standard solution	
Nitrates	≤ 0.2ppm	
Heavy Metals	≤ 0.1ppm	
Ammonium	≤ 0.2ppm	
Residue on evaporation	≤0.003%	
Oxidizable Substances	The solution remains faintly pink	
TDS	≤ 50ppm	

Time:

Parameters	Limits	Results
Description	Colourless, odourless, tasteless	
pH	5.5 – 7.5	
Conductivity	≤ 5.1µs/cm @ 25^{0}c	
Acidity/Alkalinity	NMT 0.5mL of 0.01M NaOH/HCl is required to change the solution to yellow/red.	
Total Hardness	NMT 5ppm	
Chlorides	≤ 0.2ppm	
Sulphates	The turbidity produced in test solution for at least 1 hour should not be greater than standard solution	
Nitrates	≤0.2ppm	
Heavy Metals	≤ 0.1ppm	
Ammonium	≤ 0.2ppm	
Residue on evaporation	≤0.003%	
Oxidizable Substances	The solution remains faintly pink	
TDS	≤ 50ppm	

Chemist

QC Head

CONTROL COPY