MASTER COPY

biomedical Limited	BIOMEDICAL LIMITED, NO 1 OHIMEGE ROAD, INDUSTRIAL ESTATE, ILORIN KWARA STATE, PMB 1499 PVC POUCH CERTIFICATE OF ANALYSIS (100ml) **BML/QA/STP/PKF/008/F02**

Material Name: PVC Pouch (100ml)	Batch No:	Pharmacopeia Ref: BP 2023/In-House
Manufacturer:	Supplier:	Logbook/Revision Number
MF. Date:	EX. Date:	Logbook page No.:
Sampled By:	Qty Received:	Logbook Volume:
Product Code: PKF008	No. of containers:	Analysed by:
QC. AR. No: PMF/AR/QC/ /P0	Test Completed Date:	SOP Ref. No: BML-QDG-QA-105

S/No	Test	Specification	Result
1.	Description	A clear transparent, thick, flexible airtight PVC Pouch, with no lump or bulging.	
2.	Identification (% Ash Content)	Shall comply identification tests ($A_1 \leq 5.0$)	
3.	Acidity or Alkalinity	NMT 0.8ml of 0.01M HCL VS is required	
4.	Absorbance	NMT 0.2	
5.	Residue on Evaporation	NMT 2mg	
6.	Reducing Substance	The difference between the titration volumes is not greater than 3.0ml.	
7.	Extractible Heavy Metals	NMT 2ppm	
8.	Air Tightness	When filled with air at 8.0psi, there should be no bubble when immersed in water	
9.	Colour Check	As per the approved ART WORK for the product	
10.	Lettering and spelling check	There should be no wrong spellings and the lettering must be legible	
11.	Edge to edge Length	12.80 – 13.20cm	
12.	Edge to Edge Breadth	8.50 – 9. 20cm	
13.	Neck Tube Length	4.50 – 5.50 cm	
14.	Flying wing Neck Length	4.50 – 5.50 cm	
15.	Ink Lifting Test	No ink must be lifted with Tape and when Autoclaved there should be no ink lifting or smulging	
16.	Welding Joint Length	2.00 – 2.50cm	

CONTROL COPY

17.	Neck Tube Diameter	0.75- 0.85cm	
18.	Welding Joint thickness	1.00- 1.40mm	
19.	Leakage Test	There should not be any form of leakages. Should be able to withstand temperature up to 122°C and 2.5 Bar pressure	
20.	Fill Volume Test	100 – 105ml	
21.	PVC Weight	15.0 – 15.50g	
22.	PVC Thickness	0.80 – 1.0mm	
23.	Total Bacteria Count	Must be sterile	
24.	Total Yeast &Moulds Count	Must be sterile	
25.	Pathogens	Must be absent	
26.	Pyrogen	Pyrogen Free	

CHECKER's VERDICT (box option) : **RELEASED / FAILED / HOLD / REPEAT**

Remark:

Analyst(chemist):	Analyst(microbiologist):	Reviewed By: QCH	Approved By: QAM
Sign:	Sign:	Sign:	Sign:
Date:	Date:	Date:	Date:

CONTROL COPY