	BIOMEDICAL LIMITED 1, Ohimege Road, Ilorin, Kwara State.		
	TITLE	INFUSION STABILITY CHAMBER CLEANING & INSPECTION CHECKLIST.	
	SOP REF. NO	BML-ENG-NG-003	**FORM NUMBER**	BML/ENG/NG/003/G0052
	REVIEW NUMBER	00	**REVIEW DATE**	NA
	FACTORY	Infusion	**REASON FOR REVIEW**	NA

MASTER COPY

Person In Charge: **Equipment:** **Month:**

S/N	component	Freq	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	24	25	26	27	28	29	30	31
1	Stability chamber Assembly (external Cleaning & Inspection)	D																															
2	Temperature display (btw 38-42 degree Celsius)	D																															
3	Relative humidity display (between 70-80%)	D																															
4	Temperature sensor inspection	M																															
5	Relative humidity sensor inspection	M																															
6	Backup power supply	M																															
7	Compressor assembly	M																															
8	HMI (clean and check display)	M																															
9	Circuit breaker	M																															
10	Water (replacement & bucket cleaning)	M																															
11	Stability chamber Assembly (internal Cleaning & Inspection)	M																															
12	Water hose inspection for leakages	M																															
13	Condenser filter cleaning	M																															
14	Condenser assembly Inspection	M																															
15	Temperature probes calibration	A																															
16	Relative humidity probes calibration	A																															

CONTROL COPY

Frequency

S = Shiftly **D = Daily** **W = Weekly** **F = Fortnightly**

M = Monthly **Q = Quarterly** **SA = Semi-Annually** **A = Annually**

Schedule

☑ **= Inspection Okay** R **=Reschedule Inspection**

☒ **= Defect In Inspection**

Date	Components	Defects	Action plan