	Biomedical Limited, No 1 Ohimege Road, Industrial Estate, Ilorin Kwara State, PMB 1449			MASTER COPY STAMP HERE
	Title: **SOP for Cleaning and Operation of De Lama Autoclave**			
Department	Unit:	**BML-PDF-PR-017**	**Revision No.:**	06
			Effective Date:	**April, 2025**
PRODUCTION	PROCESSING		**Next Review Date:**	**March, 2027**
Prepared By:		**Checked By:**	**Approved By:**	
Sign /date		**Sign /date**	**Sign /date**	

1.0 OBJECTIVE:

To provide the instructions for cleaning and operation of the De Lama autoclave.

2.0 SCOPE:

To provide written guidelines for cleaning and operation of De Lama Autoclave.

3.0 RESPONSIBILITIES/ACCOUNTABILITIES:

Production Pharmacist, Production Supervisors and Operators.

4.0 ABBREVIATION AND DEFINITIONS

NIL

5.0 PROCEDURES

Cleaning Procedure:

- Ensure that all cleaning equipment are clean i.e. buckets and soft cloth.
- Fill all buckets needed with water then spray the soap (if using powdered soap) into the autoclave chamber or add bar soap to soft cloth (if using bar soap) and clean.
- Then use the water to rinse the soap solution from the autoclave and drained.
- Make sure the outside part of the autoclave is clean with dry cloth provided.
- Then wash and dry all the cleaning equipment.

NOTE:
Do not use detergent in cleaning the autoclave.

6.0 Operation Procedure:

- Ensure that the boiler and compressor are on and working properly.

BIOMEDICAL LTD

- Switch on the autoclave at the control panel and press the EpW1 valve to let in the demineralized water manually.
- Check the power voltage to ensure it is at 380 volts.
- Remove the identification tag 'unsterilized' placed on each of the trolleys.

- Make sure a one filled pouch of the batch to be sterilized is fixed on the probes then wheel the three trolleys into the autoclave.
- Place the probes at the middle of the first trolley and the middle of the last trolley then lock both doors of the autoclave properly.
- Switch on the programmer and select the cycle program to be used i.e. Program 1, Program 2 and Program 3.
- Press the start button to commence sterilization, monitor the operation and record in the autoclave log sheet book.
- Monitor the temperature and pressure readings and record it in the autoclave log sheet book.
- When the temperature reaches 116°C; pressure, 1.5 Bar; allow to sterilize for 40 minutes, then allow to air to cool and finally water to complete the cooling phase.
- At the end of the sterilization cycle when the alarm rings, press end cycle alarm button to stop the alarm.
- Open the autoclave door and carefully remove the two probes then wheel out the three sterilized trolleys to the offloading section.
- Set aside the Quality Control samples for analysis.

BIOMEDICAL LTD

Materials, Equipment, Apparatus and Reagents

- PPE

6.0 REFERENCE AND ENCLOSURE

User Manual, In House Method

7.0 TRAINING INVOLVEMENT

Production manager	Production pharmacist
Sectional heads	All production staff

8.0 DISTRIBUTION AND CONTROL:

MASTER COPIES	DOCUMENTATION UNIT
CONTROLLED COPIES	PRODUCTION DEPARTMENT

9.0 CHANGE HISTORY

CURRENT REVISION	SUPERSEDE	REASON FOR REVIEW
00	Nil	First issue
06	05	• Change in sop format. • Inclusion of traceability of enclosures to parent SOP

BIOMEDICAL LTD

BIOMEDICAL LTD