EC Certificate

Directive 93/42/EEC Annex II, excluding Section 4
Full Quality Assurance System
Medical Devices

Registration No.: HD 60112658 0001

Report No.: 31691737 001

Manufacturer:

CTL Medical Corporation
4550 Excel Parkway, Ste. 300
Addison, TX 75001
USA

Products:

Non-sterile Spinal Fixation System
Non-sterile Anterior Cervical Plate System
Non-sterile Intervertebral Body Fusion System

Replaces Approval, Registration No.: HD 60105689 0001

Expiry Date: 2020-11-17

The Notified Body hereby declares that the requirements of Annex II, excluding section 4 of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex II, section 5 of the aforementioned directive. For placing on the market of class III devices covered by this certificate an EC design-examination certificate according to Annex II, section 4 is required.

Effective Date: 2016-10-04

Date: 2016-10-04

Notified Body

M.Sc. M. Aihara

TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.

10/020 d 04.08 ® TUV, TUEV and TUV are registered trademarks. Utilisation and application requires prior approval.