Riverside Analytical Laboratory
Accredited testing · 22 Mill Lane, Ashford AS1 4QT
Tel 01234 555 019 · lab@riverside-lab.example

Water Quality Test Report

Report number	WQ-2026-0382	**Date reported**	14 April 2026
Client	Ashford Parish Council	**Sample type**	Drinking water
Sampling point	Village hall, kitchen tap	**Date sampled**	9 April 2026

This report presents the results of the chemical and microbiological analysis of one drinking-water sample. The sample was received in good condition and analysed within 24 hours of collection.

1. Chemical analysis

Table 1: Chemical parameters

Parameter	Method	Result †	Limit	Status
Nitrate (as NO_3)	ISO 10304-1 ion chromatography	4.8 mg/l	50 mg/l	**Pass**
Ammonium (as NH_4)	ISO 11732 flow analysis	< 0.05 mg/l	0.50 mg/l	**Pass**
Chloride	ISO 10304-1 ion chromatography	18 mg/l	250 mg/l	**Pass**
Iron	ISO 11885 ICP-OES	0.31 mg/l	0.20 mg/l	**Fail**
Lead	ISO 17294-2 ICP-MS	2.1 µg/l	10 µg/l	**Pass**
pH	ISO 10523 electrometric	7.4	6.5 – 9.5	**Pass**
Turbidity	ISO 7027 nephelometric	0.9 NTU	4 NTU	**Pass**

† Results are the mean of three replicate measurements.

2. Microbiological analysis

Table 2: Microbiological parameters

Organism	Method	Result	Limit
Escherichia coli	ISO 9308-1	0 per 100 ml	0 per 100 ml
Coliform bacteria	ISO 9308-1	0 per 100 ml	0 per 100 ml
Colony count, 22 °C	ISO 6222	12 per ml	no abnormal change

3. Conclusion

The sample meets the drinking-water standard for all parameters except iron, which exceeds the limit of 0.20 mg/l. The elevated iron content does not pose a health risk, but it can cause staining and a metallic taste.

4. Recommendations

i. Flush the kitchen tap for two minutes before drawing water for drinking.

ii. Inspect the internal pipework of the village hall for corrosion.

iii. Repeat the iron analysis on a new sample after the inspection.

5. Sampling and handling

The sample was collected by a trained sampler in a sterilised borosilicate bottle after the tap had been run for one minute. It was kept at 4 °C during transport. The results relate only to the sample as received.

This report shall not be reproduced except in full without the written approval of the laboratory.

Analyst Date: 14 April 2026	Authorised signatory Date: 14 April 2026