Meeting with Dr Dave Collins

27 October 2026 at 12:30 p.m.

Main discussion

We presented the wrist sensing and stimulation idea and asked about measurement, testing and practical feasibility.

Dr Collins recommended starting with the wrist and a simple grasping task before attempting reaching or whole-arm movement.

He suggested starting with an accelerometer because it measures the movement we want to reduce. Measuring an M-wave would be harder, especially during movement.

He questioned whether the muscle-contraction approach could make voluntary movement harder. A useful result needs to preserve hand function as well as reduce shaking.

He asked what our device would improve compared with existing products. We raised cost and adaptive feedback, but the specific advantage still needed work.

Reproducing tremor-like movement was discussed as an exploratory idea. No detailed testing protocol was agreed.

Dr Collins and Alex were open to helping with questions and future work.

Next steps

Start with wrist movement sensing and a simple task.

Check whether reduced shaking also preserves useful grasping and movement.

Compare existing products and identify a specific improvement.

Keep Dr Collins and Alex updated and ask about stimulators when choosing equipment.

Still open

The stimulator, extra electrical measurements, equipment access and test setup were undecided.

No fixed follow-up schedule or patient-testing arrangement was established.