Meeting with Dr Martin McKeown

27 October 2026 at 1:30 p.m.

Main discussion

We asked Dr McKeown to review the wrist sensing and stimulation idea. He described tremor reduction as a difficult, high-risk problem.

He asked us to clarify whether we were targeting Parkinson's tremor or essential tremor, and what would make the device useful compared with existing options.

Medication-resistant Parkinson's tremor came up as a possible target, but we did not settle the patient group.

He explained that different muscles can contribute to similar movements and that the pattern varies between people. Selective stimulation would be difficult.

He suggested investigating whether a tolerable stimulation pattern could disrupt tremor, possibly through feedback from the hand to the brain. This was a research direction to investigate.

Comfort and useful hand function matter. He said patients may reject a treatment that is more disruptive than their tremor.

Tremor varies naturally. He suggested repeated measurements and a similar-intensity comparison stimulus when evaluating an optimized pattern.

He pointed us to his focused-ultrasound work for measurement tasks and was open to updates arranged through his assistant.

Next steps

Clarify the target tremor type, task and intended benefit.

Investigate patterned stimulation and review the suggested research.

Include comfort, hand function and a credible comparison in the evaluation plan.

Arrange a follow-up when we have concrete progress.

Still open

The stimulation pattern, electrodes, acceptance targets and study protocol were undecided. Patient recruitment and testing access were not established.