Progress Report 1 Feedback

Date: October 12, 2026

Assignment score: 120/120 - Pass

Detailed rubric: 10/12 - Level 3

Strengths:

- Clear, evidence-supported problem statement, users, and setting.

- User needs are separated from the provisional design concept.

- Uncertainties, pending user feedback, and limits of rest-tremor suppression are acknowledged.

Required improvements:

- Validate and prioritize needs with patients, care partners, and clinicians.

- Keep rest-tremor performance separate from claims about benefits during daily activities.

- Test whether wrist damping can reduce tremor while preserving voluntary movement. Use comparisons to decide whether suppression is feasible or monitoring should become the primary focus.

- Determine which monitoring information patients and clinicians would find useful.

- Check that reference [11] directly supports the device-testing authorization discussion.

- Fix inconsistent header and footer page numbers.

Rubric:

- Problem context, users, and stakeholders: 4/4.

- User needs, need statement, and intended impact: 3/4.

- Written communication and evidence: 3/4.