

Acquisition Parameter

Method: ETH_HyStar_HPLC_QTOF_POS_LowMass_Loop-AS.m Acquisition Date: 24.11.2023 14:52:13 File Name: D:\Data\bmax0261xx\BMAX026193.d Operator: Louis Bertschi

Source Type ESI Ion Polarity Positive Set Nebulizer 1.6 Bar 4500 V Set Dry Heater 230 °C 10.0 l/min Active Set Capillary Focus Set End Plate Offset Scan Begin 50 m/z -500 V Set Dry Gas Scan End 1300 m/z Set Collision Cell RF 200.0 Vpp Set Divert Valve Source







Evaluation Spectra / Validation Formula:

#	Ion Formula	Adduct	m/z	Z	Meas. m/z	mSigma	N-Rule	err [mDa]	err [ppm]
1	C26H34F7NO3PPdS	M+H	710.0915	1+	710.0929	21.8	ok	-0.5	-0.7

Calibration Info:				Mass List:					
Internal calibra	tion			#	m/z	Res.	S/N	۱%	FWHM
Date:	24.11.20	23 14:54:50	0	"	101.0085	16702	581.1	8.0	0.0060
Polarity:	Positive			2	115.0336	17721	146.7	2.1	0.0065
Calibration spec		35-1.55min,	#52-88, #208-231:	3	118.0863	20494	1185.4	17.2	0.0058
Scan				4	125.0360	16425	173.8	2.6	0.0076
Reference mass		. ,	ESI-TOF Spezial	5	147.1295	20998	318.9	4.9	0.0070
Calibration mod		ed Quadration	C	6	158.9638	20132	648.3	9.9	0.0079
Standard deviat	ion: 0.751 pp	om		7	167.0780	15845	178.4	2.7	0.0105
D (D 11: /			8	195.1737	17894	133.9	2.0	0.0109
Reference m/z	Resulting m/z	Intensity	Error [ppm]	9	203.1427	17046	135.4	2.0	0.0119
118.0863	118.0863	28138	-0.019	10	203.1920	35976	6885.5	100.0	0.0056
322.0481	322.0482	6756	0.361	11	204.1958	22516	737.4	10.7	0.0091
622.0290	622.0289	41113	-0.117	12	209.1900	21690	1312.4	18.9	0.0096
922.0098	922.0100	16617	0.256	13	210.1930	19488	212.7	3.1	0.0108
1221.9906 1521.9715	1221.9900	2543	-0.537	14	219.0238	16728	253.7	3.6	0.0131
1821.9523				15	219.1867	19664	464.1	6.6	0.0111
2121.9332				16	223.2055	22954	1363.8	19.2	0.0097
2421.9140				17	224.2088	17003	197.0	2.8	0.0132
2721.8948				18	227.2366	18224	374.0	5.2	0.0125
2721.0540				19	237.2211	22325	1007.9	13.7	0.0106
				20	238.2243	18575	172.5	2.3	0.0128
				21	241.1689	18089	159.6	2.2	0.0133
				22	249.2202	17937	196.2	2.6	0.0139
				23	251.2368	21611	840.9	11.2	0.0116
				24	252.2399	18196	150.9	2.0	0.0139
				25	255.2681	21042	480.8	6.4	0.0121
				26	277.2513	17593	167.2	2.2	0.0158
				27	293.2835	19146	304.3	4.0	0.0153
				28	295.2988	17455	161.4	2.1	0.0169
				29	307.2993	17356	168.1	2.2	0.0177
				30	321.3148	19922	257.2	3.4	0.0161
				31	322.0482	19246	313.4	4.1	0.0167
				32	338.3416	18947	145.8	1.9	0.0179
				33 34	379.2593 380.2623	20393 17550	872.0 197.7	11.5 2.6	0.0186 0.0217
				35	411.2853	17350	192.2	2.5	0.0236
				36	430.9138	18841	312.2	4.0	0.0230
				37	622.0289	30425	2876.5	25.1	0.0204
				38	623.0324	19139	421.3	3.7	0.0326
				39	922.0100	23270	1795.1	10.1	0.0396
				40	923.0129	17679	356.1	2.0	0.0522
				#			S/N	I %	
					m/z	Res.	3/N		FWHM 0.0365
				1	706.0936	19335		2.9	0.0365
				2 3	707.0968 708.0920	19363 19390		0.9 32.1	0.0365
				3 4	708.0920	19390		73.3	0.0365 0.0365
				5	710.0933	19417		100.0	0.0365
				6	710.0924	19443		29.8	0.0365
				7	711.0940	19472		84.1	0.0365
				8	713.0950	19527		24.1	0.0365
				U	7 10.0000	10021		∠ ¬. I	5.0000

BMAX026193 Michael Bogdos/Morandi - F4-NTf-OMe-PtBu3 - DCM/MeOH - --



Eidgenössische Technische Hochschule Zürich Swiss Federal Institute of Technology Zurich

#	m/z	Res.	S/N	I %	FWHM
9	714.0931	19554		40.8	0.0365
10	715.0961	19582		11.4	0.0365
11	716.0944	19609		3.3	0.0365
12	717.0954	19636		0.7	0.0365