

Acquisition Parameter

Method: ${\sf ETH_HyStar_HPLC_QTOF_POS_LowMass_Loop\text{-}AS.m}$ Acquisition Date: 25.03.2024 08:44:44 File Name: D:\Data\bmax0276xx\BMAX027672.d Operator: Michael Meier

Set Nebulizer Set Dry Heater Ion Polarity Source Type ESI Positive 1.6 Bar 4500 V Focus Active Set Capillary

230 °C 10.0 l/min -500 V Scan Begin 50 m/z Set End Plate Offset Set Dry Gas Scan End 1300 m/z Set Collision Cell RF 200.0 Vpp Set Divert Valve Source







Evaluation Spectra / Validation Formula:

Calibration Info:				Mass List:					
Internal calibration				#		Dag	S/N	۱%	FWHM
Date:	25.03.20	024 10:33:5	1		m/z 209.1074	Res.			
Polarity: Positive			1		21358	576.9	29.2	0.0098	
Calibration spectrum: +MS, 0.41-1.46min, #60-80, #192-217:			2	209.1901 223.2057	20286 19164	373.8 335.7	18.9 16.9	0.0103	
Scan				3	223.2057	20258	275.0	13.8	0.0116
Reference mass list: ESI: Tunemix (pos) ESI-TOF Spezial				4					0.0117
Calibration mode: Enhanced Quadratic				5	251.2368 444.5698	19254	229.3 427.1	11.5	0.0130
Standard deviation: 0.438 ppm				6		22654		21.8	0.0196
• •				7	445.0695	24617	607.4	31.0	0.0181
Reference m/z	Resulting m/z	Intensity	Error [ppm]	8	445.5708	19060	260.4	13.3	0.0234
118.0863	118.0863	5516	0.026	9	446.0695	21475	440.4	22.5	0.0208
322.0481	322.0481	4433	-0.142	10	446.5710	19463	234.6	11.9	0.0229
622.0290	622.0290	10215	0.120	11	447.0702	19340	229.3	11.7	0.0231
922.0098	922.0096	5103	-0.212	12	452.0666	21592	528.6	26.7	0.0209
1221.9906	1221.9910	1155	0.299	13	452.5669	30935	1379.7	69.7	0.0146
1521.9715				14 15	453.0665	33024	1980.8 920.7	100.0 46.4	0.0137
1821.9523					453.5682	26169			0.0173
2121.9332				16	454.0664	30288	1441.1	72.7	0.0150
2421.9140				17	454.5681	24635	750.2	37.9	0.0185
2721.8948				18	455.0676	24224	726.2	36.6	0.0188
				19	455.5691	20591	356.6	18.0	0.0221
				20	491.0277	17594	215.7	10.6	0.0279
				21	497.5221	20041	233.5	11.4	0.0248
				22	498.0220	19896	256.2	12.5	0.0250
				23	498.5223 499.0219	20154	254.6	12.4	0.0247
				24		19983	292.0	14.2	0.0250
				25	499.5228	18675	192.1	9.3	0.0267
				26 27	500.0223 524.9921	18653 19045	214.4 219.8	10.4 10.6	0.0268 0.0276
							219.6		
				28	526.9914	18885		10.7	0.0279
				29 30	535.0226 537.0226	19218 18672	237.3 206.4	11.4 9.9	0.0278 0.0288
				31	583.0576	19507	234.6	11.1	0.0299
				32	622.0290	21381	314.1	14.7	0.0291
				33	697.0798	19083	212.9	9.4	0.0365
				34	697.5798	19166	285.0	12.6	0.0364
				35	698.0797	19453	334.8	14.8	0.0359
				36	698.5799	20287	345.0	15.3	0.0344
				37	699.0795	20149	383.0	16.9	0.0347
				38	699.5804	19753	312.1	13.8	0.0354
				39	700.0801	19957	311.5	13.8	0.0351
				40	905.1253	19186	301.7	9.7	0.0472