
Acquisition Parameter


Method: ETH_HyStar_HPLC_QTOF_POS_MidMass_Loop-AS.m Acquisition Date: 25.03.2024 07:43:52

File Name: D:\Data\hmax0276xy\BMAX027681 d Operator: Michael Meier

File Name: D:\Data\bmax0276xx\BMAX027681.d Operator: Michael Meier

Source Type ESI Ion Polarity Positive Set Nebulizer 1.6 Bar 4500 V Set Dry Heater 230 °C 10.0 l/min Active Set Capillary Focus -500 V Scan Begin 100 m/z Set End Plate Offset Set Dry Gas Scan End 3000 m/z Set Collision Cell RF 1500.0 Vpp Set Divert Valve Source







Evaluation Spectra / Validation Formula:

#	Ion Formula	Adduct	m/z	Z	Meas. m/z	mSigma	N-Rule	err [mDa]	err [ppm]
1	C50H37CIF4NO3P2PdS	M+H	1010.0623	1+	1010.0609	26.9	ok	2.8	2.8

Calibration Info:				Mass	Mass List:					
Internal calibrat	tion			#	m/z	Res.	S/N	۱%	FWHM	
Date:	25.03.20	024 07:48:23	3	1	337.1153	25791	501.8		0.0131	
Polarity:	Positive					21397		7.5		
Calibration spectrum: +MS, 0.60-1.16min, #89-172: Scan			#89-172: Scan	2	338.1184		118.5	1.8	0.0158	
Reference mass	list: ESI: Tur	nemix (pos)	ESI-TOF Spezial	3	339.1299	18817	126.2	1.9	0.0180	
Calibration mode	e: Quadrat	ic,	·	4	622.0290	43073	3890.2	39.4	0.0144	
Standard deviati	on: 0.412 pp	om		5	623.0331	27000	477.2	4.8	0.0231	
				6	663.0399	24316	175.9	1.7	0.0273	
Reference m/z	Resulting m/z	Intensity	Error [ppm]	7	663.4547	22068	222.5	2.1	0.0301	
118.0863		,	.,, .	8	693.9973	27694	249.7	2.3	0.0251	
322.0481				9	694.9988	29986	575.9	5.2	0.0232	
622.0290	622.0290	64855	0.112	10	695.9974	33640	926.4	8.4	0.0207	
922.0098	922.0095	46001	-0.289	11	696.9984	27626	433.4	3.9	0.0252	
1221.9906	1221.9909	19116	0.217	12	697.9964	34009	968.6	8.7	0.0205	
1521.9715				13	698.9991	23955	264.5	2.4	0.0292	
1821.9523	1821.9522	68	-0.037	14	699.9969	29340	531.3	4.8	0.0239	
2121.9332	1021.0022	00	0.007	15	922.0095	43391	3912.2	27.9	0.0212	
2421.9140				16	923.0141	28953	771.1	5.5	0.0319	
2721.8948				17	1006.0659	25941	279.7	1.9	0.0388	
2721.0010				18	1008.0623	48840	3725.3	24.9	0.0206	
				19	1009.0621	47928	8154.8	54.5	0.0211	
				20	1010.0609	51751	14203.2	94.9	0.0195	
				21	1011.0630	51970	9138.1	61.0	0.0195	
				22	1012.0600	54392	14972.4	100.0	0.0186	
				23	1013.0639	51884	7189.4	48.0	0.0195	
				24	1014.0611	53635	9025.5	60.2	0.0189	
				25	1015.0652	46242	3658.6	24.4	0.0220	
				26	1016.0642	43671	2260.5	15.1	0.0233	
				27	1017.0667	32544	857.1	5.7	0.0313	
				28	1030.0460	26519	511.1	3.4	0.0388	
				29	1031.0466	33387	1105.6	7.3	0.0309	
				30	1032.0457	37130	1599.2	10.5	0.0278	
				31	1033.0473	29061	940.2	6.2	0.0355	
				32	1034.0451	38323	1564.1	10.3	0.0270	
				33	1035.0483	28344	749.6	4.9	0.0365	
				34	1036.0459	32436	986.2	6.5	0.0319	
				35	1037.0470	24090	425.9	2.8	0.0430	
				36	1038.0429	19645	263.0	1.7	0.0528	
				37	1048.0223	23595	243.9	1.6	0.0444	
				38	1050.0209	27570	270.8	1.8	0.0381	
				39	1221.9909	34789	1998.6	11.6	0.0351	
				40	1222.9954	25893	505.7	2.9	0.0472	
				#	m/z	Res.	S/N	I %	FWHM	
				1	1006.0645	51546		2.2	0.0195	
				2	1007.0677	51597		1.2	0.0195	
				3	1008.0629	51648		24.9	0.0195	
				4	1009.0644	51700		61.4	0.0195	
				5	1010.0637	51751		97.5	0.0195	
				6	1011.0648	51802		63.1	0.0195	
				7	1012.0629	51853		100.0	0.0195	
				8	1013.0654	51905		49.3	0.0195	



Eidgenössische Technische Hochschule Zürich Swiss Federal Institute of Technology Zurich

#	m/z	Res.	S/N	I %	FWHM
9	1014.0634	51956		60.1	0.0195
10	1015.0658	52007		28.5	0.0195
11	1016.0641	52058		17.7	0.0195
12	1017.0654	52110		7.0	0.0195
13	1018.0661	52161		2.1	0.0195
14	1019.0667	52212		0.5	0.0195
15	1020.0680	52264		0.1	0.0195