

Tuneable Light Source Testing Report

Project No.:	17125
End User:	University of California, Santa Cruz
End User Country:	United States

Manufacturer	Sciencetech, Inc.
Model:	TLS-55-X300
System Voltage Specification:	100 – 240 VAC, 50/60 Hz
Serial No.:	00040043

Performed By:	RP
Date [YYYY-MM-DD]:	2024-09-20
Reviewed By:	JF
Date of Issue [YYYY-MM-DD]:	2024-09-23

Disclaimer: The values presented in this test document have been demonstrated with your test equipment at Sciencetech under ideal measurement and environmental conditions (except where expressly noted). Variation from these values is expected due to variables introduced during shipping and installation of the system. The equipment is expected to meet the required performance specifications after installation if the instructions supplied are properly followed, however some variation from the values reported in this report is normal and expected. In order to verify and calibrate your equipment for values such as absolute irradiance or sun value it is recommended that a calibrated reference cell be used.

Measurements	Value	Units	Notes
Power at 0 nm* Grating 1:	2.2	W	6 mm slits Measured with XPL12S SN: 243780
Power at 0 nm Grating 2:	1.1	W	6 mm slits Measured with XPL12S SN: 243780
Power at 0 nm Grating 3:	0.35	W	6 mm slits Measured with XPL12S SN: 243780
Power at 400 nm Grating 1:	15.9	mW	6 mm slits Measured with PH100-SiUV SN:237981
Power at 800 nm Grating 2:	16.3	mW	6 mm slits Measured with PH100-SiUV SN:237981
Power at 400 nm Grating 1:	7.4	mW	1 mm slits Measured with PH100-SiUV SN:237981
Power at 800 nm Grating 2:	0.3	mW	1 mm slits Measured with PH100-SiUV SN:237981
Measurement Configuration:	Focus collimated beam via F/1 lens onto detector surface		
Measurement Date (YYYY-MM-DD):	2023-09-20		

*0nm is the configuration in which there is total specular reflection from the grating surface.