



Star Nuke Consulting Engineering Services Private Limited

AERB (Govt. of India) Authorized - ISO 9001-2015 Certified company

Portable X - Ray Machine (V2) (Q2)

COMPLIANCE SHEET

GEM/2025/B/7007849

Dated: 03-02-2026

Sr. No.	Specification	Compliance
1.	General Description:*	
•	A high-frequency mobile X-ray system designed for bedside and emergency imaging, equipped with a spring-balanced arm and compact maneuverable design for indoor mobility.	YES
•	The unit shall be available in 3.5 kW, 4.2 kW, or 5.2 kW generator configurations.	YES
2.	X-Ray Generator:*	
•	Type: High-frequency X-ray generator	YES
•	Output Power Options: 3.5 kW / 4.2 kW / 5.2 kW	YES
•	X-ray Ripple Frequency: 100 kHz	YES
•	Input Power: Single phase, 220V, 50 Hz $\pm 10\%$, 15A	YES
•	Output Voltage Range: 40 – 125 kV	YES
•	mA Range: 40 – 100 mA	YES
•	Maximum mAs: 200 mAs	YES
3.	X-Ray Tube Assembly:*	
•	Tube Type: Stationary or Rotating anode (as per configuration)	YES
•	Focal Spot Size: 1.5 mm	YES
•	Tube Mount: Spring-balanced arm with folding feature for compact storage and easy mobility	YES
•	Tube Head Movement: 360° rotation with locking mechanism for precise positioning	YES
4.	Mobile Stand & Movement:*	
•	Design: Compact, lightweight with spring-balanced arm	YES
•	Hand Braking System for safe parking	YES
•	Easy movement on slopes, inclines, and ramps	YES
•	Big rear wheels and ergonomic handles	YES
•	Folding arm for space-saving and secure transport	YES
5.	Control Panel:*	
•	Type: Microprocessor-based digital feather-touch control panel	YES
•	Display: Integrated 14"/15" touchscreen LCD mounted on mobile stand	YES
•	APR (Anatomical Programmed Radiography): Yes, with body part and patient type presets	YES
•	Programmable Anatomical Technique: Adjustable by operator	YES
6.	Optional DR Upgrade:*	
•	Flat Panel Detector Compatibility: Upgradeable to Digital Radiography	YES
•	Panel Sizes Supported: 14" x 17" and 17" x 17"	YES
•	Wireless Transmission to Laptop	YES
•	Detector Type: Konica Minolta AeroDR or equivalent	YES
•	Pixel Size: 150 microns (NS or GS type)	YES



Star Nuke Consulting Engineering Services Private Limited

AERB (Govt. of India) Authorized - ISO 9001-2015 Certified company

7.	Safety & Compliance:*	
•	Radiation safety with collimator light and exposure controls	YES
•	Compliance: IEC and BIS/CE certification (as applicable)	YES
•	AERB type approval (mandatory in India)	YES
•	Emergency stop button and overload protection circuit	YES
8.	Accessories (Mandatory):*	
•	Cassette storage compartment	YES
•	Power cord and charger	YES
•	Collimator with light source	YES
•	Operator and service manuals	YES
•	Optional wireless detector and laptop for DR use	YES

Balram Singh Rajora
Director & CEO
ceo@starnukemail.net

