



**भारतीय मानक ब्यूरो**  
**BUREAU OF INDIAN STANDARDS**

जयपुर शाखा कार्यालय-1  
**Jaipur Branch Office-I**

Address: Manak Bhawan  
Prithviraj Road, C-Scheme- Jaipur| 302005

दूरभाष Phones 0141-2223282,2223283,2223284

फैक्स Fax 0141-2223286

ई-मेल e-mail: jpbo@bis.gov.in

वेब web : <http://www.bis.org.in>

हमारा संदर्भ: JBO/एल-8400193016

दिनांक 05/04/2022

**M/s Star Nuke Consulting Engineering Services Pvt Ltd,**  
**Plot No 25, Navkar Nagar, Kalyanpura, New Sanganer Road,**  
**Sanganer, Distt.:Jaipur (Rajasthan)-302020**

महोदय(यों),

भा मा ब्यूरो प्रमाणन मुहर लाइसेंस संख्या सीएम/एल-8400193016 के संदर्भ में है जिसे आपने Diagnostic Medical X-ray Equipment - Part 1 : General and Safety Requirements IS 7620 (Pt.1) : 1986 के अनुसार है, जो 10/01/2023 तक वैध है।

हमें आपको यह सूचित करते हुए हर्ष है कि भा मा ब्यूरो ने आपके लाइसेंस में दिनांक 07-03-2022 से नई किस्म समाविष्ट करने के लिए सहमति दे दी है।।:

जब और जहाँ नई किस्म के मुहरांकन को शुरू करते हैं, कृपया हमें सूचित करें। आपको सलाह दी जाती है कि उक्त किस्म का एसआईटी के अनुसार परीक्षण करने के उपरांत अलग से रिकार्ड रखें।

नई किस्म के लिए पृष्ठांकन शीट संख्या 1 संलग्न है।

सधन्यवाद,

भवदीय,  
**कनिका**  
(कनिका कालिया)  
वैज्ञानिक 'ई' एवं प्रमुख



# भारतीय मानक ब्यूरो BUREAU OF INDIAN STANDARDS

उपभोक्ता मामले, खाद्य एवं सार्वजनिक वितरण मंत्रालय,  
भारत सरकार

Ministry of Consumer Affairs, Food and Public  
Distribution, GOVT. OF INDIA

जयपुर शाखा कार्यालय-1

**Jaipur Branch Office-I**

पता/Address: मानक भवन, पृथ्वीराज रोड,

सी-स्कीम, जयपुर-302005

Manak Bhawan, Prithviraj Road

C-Scheme- Jaipur 302005

दूरभाष/Phones 0141-2223282, 2223283, 2223284

फैक्स/fax: 0141-2223281

ई-मेल/E-Mail: jpbo@bis.gov.in

वेबसाइट/web : <http://www.bis.org.in>

लाइसेंस संख्या सी एम/एल- 8400193016 की संलग्नी  
ATTACHMENT TO LICENCE NO. CM/L- 8400193016

| सी एम/एल<br>संख्या<br>CML/NO | लाइसेंसधारी का पता सहित नाम<br>NAME OF THE LICENSEE WITH THE<br>ADDRESS                                                                                                       | उत्पाद का नाम<br>NAME OF THE PRODUCT                                                   | भा मा<br>संख्या<br>IS NO.   |
|------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|----------------------------------------------------------------------------------------|-----------------------------|
| 8400193016                   | M/s Star Nuke Consulting Engineering<br>Services Pvt Ltd,<br>Plot No 25, Navkar Nagar, ,<br>Kalyanpura, New Sanganer Road,<br>Sanganer, Distt.:Jaipur (Rajasthan)-<br>302020, | Diagnostic Medical X-ray<br>Equipment - Part 1 :<br>General and Safety<br>Requirements | IS 7620 :<br>PART<br>1:1986 |

पृष्ठांकन संख्या 1 दिनांक Dated: 05/04/22  
ENDORSEMENT NO. 1 DATED

निम्न अतिरिक्त (आकारों/प्रकारों/श्रेणियों, आदि) लाइसेंस के प्रथम अनुसूची के कालम (2) में और द्वितीय अनुसूची के कालम (1) में प्रथम अनुसूची के कालम (1) में मानक मुहर सहित दिनांक 07-03-2022 से प्रभावी होने के साथ सम्मिलित किया जाता है।

The following additional (sizes/types/grades etc.) has (have) been included in Columns(2) of the first schedule and column (1) of the the Second Schedule of the Licence alongwith the Standard Mark in Column (1) of First Schedule with effect from 07-03-2022

1. Diagnostic X-ray machine Permanently Installed type, Six Pulse Line Frequency Generator, For rated line voltage 230 + 10 V, frequency 50 Hz, Three phase, Current 300 mA, Rated Peak Output Voltage 125 kVp, Output power 24 kW with Horizontal Bucky Table,
2. Diagnostic X-ray machine Stationary type, High frequency generator, For rated line voltage 230 V, frequency 50 Hz, Single phase, Output Current 100 mA, Rated Peak Output Voltage 110 kV, Output power 5 KW with Horizontal Bucky Table.
3. Diagnostic X-ray machine Mobile type, High frequency generator, For rated line voltage 230 V, frequency 50 Hz, Single phase, Output Current 100 mA, Rated Peak Output Voltage 110 kV, Output power 5 KW .
4. Diagnostic X-ray machine Stationary type, High frequency generator, For rated line voltage 230 V, frequency 50 Hz, Single phase, Output Current 320 mA, Rated Peak Output Voltage 125 kV, Output power 20 KW with Horizontal Bucky Table.
5. Diagnostic X-ray machine Portable Handheld Dental, For rated line voltage 240 V, 50 Hz Single Phase, Current 2.5 mA, Rated Peak voltage 65 kV, Output Power 162.5 W High Frequency Generator

लाइसेंस की अन्य निबन्धन व शर्तें यथावत हैं।

Other terms and conditions of Licence remain the same.

कनिका  
(कनिका कालिया)  
वैज्ञानिक 'ई' एवं प्रमुख